

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080816\
 Data File : BE092218.D
 Acq On : 8 Aug 2016 15:45
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 09 02:41:46 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 08 18:55:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.680	0.696	-2.4	87	0.00
3	n-Nitrosodimethylamine	0.942	0.741	21.3	86	0.00
4 S	2-Fluorophenol	1.284	1.100	14.3	85	0.01
5 S	Phenol-d6	1.711	1.456	14.9	81	0.01
6 I	Naphthalene-d8	1.000	1.000	0.0	93	-0.02
7 S	Nitrobenzene-d5	0.395	0.318	19.5	79	0.00
8	Naphthalene	1.159	1.114	3.9	93	0.00
9	2-Methylnaphthalene	0.743	0.701	5.7	94	0.00
10 I	Acenaphthene-d10	1.000	1.000	0.0	94	-0.02
11 S	2,4,6-Tribromophenol	0.162	0.135	16.7	94	0.00
12 S	2-Fluorobiphenyl	1.584	1.542	2.7	96	0.00
13	Acenaphthylene	8.821	8.452	4.2	96	0.00
14	Acenaphthene	1.406	1.386	1.4	96	0.00
15	Fluorene	1.783	1.708	4.2	96	-0.01
16 I	Phenanthrene-d10	1.000	1.000	0.0	99	-0.02
17	Hexachlorobenzene	0.245	0.228	6.9	92	0.00
18	Pentachlorophenol	0.068	0.068	0.0	122	0.00
19	Phenanthrene	1.460	1.378	5.6	99	0.00
20	Anthracene	1.331	1.220	8.3	97	0.00
21	Fluoranthene	6.116	5.570	8.9	94	0.00
22 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
23	Pyrene	5.228	5.197	0.6	95	0.00
24 S	Terphenyl-d14	0.718	0.727	-1.3	97	-0.02
25	Benzo(a)anthracene	5.496	5.202	5.3	89	0.00
26	Chrysene	5.505	5.330	3.2	90	0.00
27	Indeno(1,2,3-cd)pyrene	5.978	5.245	12.3	84	0.00
28 I	Perylene-d12	1.000	1.000	0.0	85	0.00
29	Benzo(b)fluoranthene	1.368	1.248	8.8	76	0.00
30	Benzo(k)fluoranthene	1.369	1.363	0.4	97	-0.01
31 C	Benzo(a)pyrene	1.257	1.220	2.9	86	0.00
32	Dibenzo(a,h)anthracene	1.200	1.091	9.1	82	0.00
33	Benzo(g,h,i)perylene	5.122	4.785	6.6	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0