

Data Path : Z:\HPCHEM1\BNA_E\Data\BE080816\
 Data File : BE092232.D
 Acq On : 9 Aug 2016 1:14
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 09 02:44:37 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-PAH-BE080216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 08 18:55:30 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	87	0.00
2	1,4-Dioxane	0.400	0.394	1.5	83	0.00
3	n-Nitrosodimethylamine	0.400	0.407	-1.7	87	0.00
4 S	2-Fluorophenol	0.400	0.348	13.0	86	0.01
5 S	Phenol-d6	0.400	0.348	13.0	82	0.01
6 I	Naphthalene-d8	5.000	5.000	0.0	93	-0.01
7 S	Nitrobenzene-d5	0.400	0.326	18.5	80	0.00
8	Naphthalene	0.400	0.404	-1.0	92	0.00
9	2-Methylnaphthalene	0.400	0.406	-1.5	94	0.00
10 I	Acenaphthene-d10	5.000	5.000	0.0	94	-0.02
11 S	2,4,6-Tribromophenol	0.400	0.338	15.5	95	0.00
12 S	2-Fluorobiphenyl	0.400	0.386	3.5	95	0.00
13	Acenaphthylene	0.400	0.379	5.3	95	-0.02
14	Acenaphthene	0.400	0.388	3.0	94	0.00
15	Fluorene	0.400	0.373	6.8	93	-0.01
16 I	Phenanthrene-d10	5.000	5.000	0.0	95	-0.02
17	Hexachlorobenzene	0.400	0.375	6.3	88	0.00
18	Pentachlorophenol	0.400	0.420	-5.0	91	0.00
19	Phenanthrene	0.400	0.380	5.0	96	0.00
20	Anthracene	0.400	0.366	8.5	92	0.00
21	Fluoranthene	0.400	0.368	8.0	91	0.00
22 I	Chrysene-d12	5.000	5.000	0.0	83	0.00
23	Pyrene	0.400	0.400	0.0	90	0.00
24 S	Terphenyl-d14	0.400	0.420	-5.0	95	-0.02
25	Benzo(a)anthracene	0.400	0.400	0.0	89	0.00
26	Chrysene	0.400	0.391	2.3	86	0.00
27	Indeno(1,2,3-cd)pyrene	0.400	0.365	8.8	82	0.00
28 I	Perylene-d12	5.000	5.000	0.0	82	0.00
29	Benzo(b)fluoranthene	0.400	0.376	6.0	76	0.00
30	Benzo(k)fluoranthene	0.400	0.389	2.8	92	0.00
31 C	Benzo(a)pyrene	0.400	0.393	1.8	84	0.00
32	Dibenzo(a,h)anthracene	0.400	0.375	6.3	82	0.00
33	Benzo(g,h,i)perylene	0.400	0.376	6.0	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0