

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080817\  
 Data File : BE093711.D  
 Acq On : 8 Aug 2017 11:14  
 Operator : SJ/JU  
 Sample : PB101071BL  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 SBLK71

Quant Time: Aug 09 04:21:26 2017  
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE072417.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Aug 09 04:20:36 2017  
 Response via : Initial Calibration


| Internal Standards          | R.T.  | QIon | Response | Conc  | Units  | Dev(Min)    |
|-----------------------------|-------|------|----------|-------|--------|-------------|
| 1) 1,4-Dichlorobenzene-d4   | 7.92  | 152  | 3366     | 0.40  | ng/ul  | 0.00        |
| 2) Naphthalene-d8           | 10.71 | 136  | 17086    | 0.40  | ng/ul  | 0.00        |
| 6) Acenaphthene-d10         | 14.53 | 164  | 10509    | 0.40  | ng/ul  | 0.00        |
| 10) Phenanthrene-d10        | 0.00  | 188  | 0        | 0.00  | ng/ul  | -17.26      |
| 16) Chrysene-d12            | 21.41 | 240  | 23362    | 0.40  | ng/ul  | 0.00        |
| 20) Perylene-d12            | 23.75 | 264  | 1535603  | 0.40  | ng/ul  | -0.16       |
| System Monitoring Compounds |       |      |          |       |        |             |
| 4) 2-Methylnaphthalene-d10  | 12.29 | 152  | 7143     | 0.26  | ng/ul  | 0.00        |
| 14) Fluoranthene-d10        | 19.28 | 212  | 17740    | 0.00  | ng/ul  | 0.00        |
| Target Compounds            |       |      |          |       |        |             |
| 17) Pyrene                  | 19.64 | 202  | 2429     | 0.036 | ng/ul# | Ovalue<br>1 |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE080817\  
Data File : BE093711.D  
Acq On : 8 Aug 2017 11:14  
Operator : SJ/JU  
Sample : PB101071BL  
Misc :  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
SBLK71

Quant Time: Aug 09 04:21:26 2017  
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE072417.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Aug 09 04:20:36 2017  
Response via : Initial Calibration





#17

Pvrene

Concen: 0.036 ng/ul

RT: 19.64 min Scan# 2826

Delta R.T. -0.02 min

Lab File: BE093711.D


Acq: 8 Aug 2017 11:14

Instrument :

BNA\_E

ClientSampleId :

SBLK71



Tot Ion:202 Resp: 2429

Ion Ratio Lower Upper

202 100

101 170.3 18.2 27.4#

100 895.7 15.6 23.4#

