

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE080817\
 Data File : BE093738.D
 Acq On : 9 Aug 2017 3:56
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.446

Quant Time: Aug 09 07:07:08 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 09 05:31:57 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
3	Naphthalene	0.973	0.996	-2.4	89	0.00
4 SURR	2-Methylnaphthalene-d10	0.648	0.645	0.5	87	0.00
5	2-Methylnaphthalene	0.710	0.703	1.0	86	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
7	Acenaphthylene	1.673	1.746	-4.4	83	0.00
8 C	Acenaphthene	1.316	1.312	0.3	78	0.00
9	Fluorene	1.664	1.534	7.8	73	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	66	0.00
11	Pentachlorophenol	0.066	0.060	9.1	67	0.00
12	Phenanthrene	1.089	1.068	1.9	67	0.00
13	Anthracene	1.061	1.021	3.8	66	0.00
14 SURR	Fluoranthene-d10	1.263	1.113	11.9	60	0.00
15 C	Fluoranthene	1.362	1.191	12.6	60	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	59	0.00
17	Pyrene	1.168	1.179	-0.9	60	0.00
18	Benzo(a)anthracene	1.160	1.125	3.0	59	0.00
19	Chrysene	1.112	1.100	1.1	59	0.00
20 I	Perylene-d12	1.000	1.000	0.0	59	0.00
21	Benzo(b)fluoranthene	1.237	1.176	4.9	58	0.00
22	Benzo(k)fluoranthene	1.220	1.209	0.9	61	0.00
23 C	Benzo(a)pyrene	1.177	1.168	0.8	61	0.00
24	Indeno(1,2,3-cd)pyrene	1.299	1.249	3.8	59	0.00
25	Dibenzo(a,h)anthracene	1.086	1.042	4.1	59	0.00
26	Benzo(g,h,i)perylene	1.095	0.995	9.1	56	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0