


1,4-Dioxane

Response Ratio




Resp Ratio = $5.53 \times 10^{-1} \times \text{Amt} + 9.06 \times 10^{-3}$
Coef of Det (r^2) = 0.999 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080916.M
Calibration Table Last Updated: Wed Aug 10 11:12:32 2016


n-Nitrosodimethylamine

Response Ratio




Resp Ratio = $9.56 \times 10^{-1} \times \text{Amt} - 3.00 \times 10^{-2}$
Coef of Det (r^2) = 0.999 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080916.M
Calibration Table Last Updated: Wed Aug 10 11:12:32 2016




Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080916.M
Calibration Table Last Updated: Wed Aug 10 11:12:32 2016



Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080916.M
Calibration Table Last Updated: Wed Aug 10 11:12:32 2016

Bis(2-ethylhexyl)phthalate

Response Ratio



Resp Ratio = 1.05e+000 * Amt - 4.91e-002
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080916.M
Calibration Table Last Updated: Wed Aug 10 11:12:32 2016