

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081015\
 Data File : BE090406.D
 Acq On : 10 Aug 2015 12:10
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 11 03:22:00 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	117	0.00
2	1,4-Dioxane	1.000	1.193	-19.3	122	0.00
3	n-Nitrosodimethylamine	1.000	0.999	0.1	118	0.00
4 S	2-Fluorophenol	1.000	1.251	-25.1#	157	0.00
5 S	Phenol-d6	1.000	1.098	-9.8	145	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	118	0.00
7 S	Nitrobenzene-d5	1.000	1.012	-1.2	130	0.00
8	Nitrobenzene	1.000	0.991	0.9	135	0.00
9	Naphthalene	1.000	1.007	-0.7	120	0.00
10	Hexachlorobutadiene	1.000	1.086	-8.6	117	0.00
11	2-Methylnaphthalene	1.000	1.059	-5.9	131	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	124	0.00
13 S	2,4,6-Tribromophenol	1.000	1.198	-19.8	170	0.00
14 S	2-Fluorobiphenyl	1.000	1.038	-3.8	127	0.00
15	Acenaphthylene	1.000	1.025	-2.5	128	0.00
16	Acenaphthene	1.000	0.998	0.2	123	0.00
17	Fluorene	1.000	1.006	-0.6	123	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	122	0.00
19	4-Bromophenyl-phenylether	1.000	1.027	-2.7	128	0.00
20	Hexachlorobenzene	1.000	0.939	6.1	114	0.00
21	Pentachlorophenol	1.000	1.039	-3.9	152	0.00
22	Phenanthrene	1.000	0.983	1.7	121	0.00
23	Anthracene	1.000	1.041	-4.1	130	-0.02
24	Fluoranthene	1.000	1.045	-4.5	132	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	126	0.00
26	Pyrene	1.000	1.030	-3.0	134	0.00
27 S	Terphenyl-d14	1.000	1.047	-4.7	133	0.00
28	Benzo(a)anthracene	1.000	1.027	-2.7	144	0.00
29	Chrysene	1.000	1.025	-2.5	126	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.996	0.4	142	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.001	-0.1	154	0.00
32 I	Perylene-d12	5.000	5.000	0.0	131	0.00
33	Benzo(b)fluoranthene	1.000	1.040	-4.0	150	0.00
34	Benzo(k)fluoranthene	1.000	0.959	4.1	122	0.00
35 C	Benzo(a)pyrene	1.000	1.022	-2.2	145	0.00
36	Dibenzo(a,h)anthracene	1.000	1.010	-1.0	159	0.00
37	Benzo(g,h,i)perylene	1.000	1.038	-3.8	140	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0