

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081015\
 Data File : BE090406.D
 Acq On : 10 Aug 2015 12:10
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 11 03:22:00 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.687	0.767	-11.6	122	0.00
3	n-Nitrosodimethylamine	0.773	0.773	0.0	118	0.00
4 S	2-Fluorophenol	0.750	0.938	-25.1#	157#	0.00
5 S	Phenol-d6	1.241	1.427	-15.0	145	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
7 S	Nitrobenzene-d5	0.331	0.344	-3.9	130	0.00
8	Nitrobenzene	0.349	0.355	-1.7	135	0.00
9	Naphthalene	1.104	1.112	-0.7	120	0.00
10	Hexachlorobutadiene	0.191	0.191	0.0	117	0.00
11	2-Methylnaphthalene	0.640	0.678	-5.9	131	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
13 S	2,4,6-Tribromophenol	0.095	0.102	-7.4	170#	0.00
14 S	2-Fluorobiphenyl	1.375	1.428	-3.9	127	0.00
15	Acenaphthylene	8.098	8.299	-2.5	128	0.00
16	Acenaphthene	1.259	1.257	0.2	123	0.00
17	Fluorene	1.635	1.645	-0.6	123	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
19	4-Bromophenyl-phenylether	0.194	0.200	-3.1	128	0.00
20	Hexachlorobenzene	1.005	0.943	6.2	114	0.00
21	Pentachlorophenol	0.044	0.034	22.7	152#	0.00
22	Phenanthrene	1.251	1.230	1.7	121	0.00
23	Anthracene	1.028	1.070	-4.1	130	-0.02
24	Fluoranthene	1.231	1.287	-4.5	132	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	126	0.00
26	Pyrene	1.084	1.116	-3.0	134	0.00
27 S	Terphenyl-d14	0.737	0.771	-4.6	133	0.00
28	Benzo(a)anthracene	1.108	1.138	-2.7	144	0.00
29	Chrysene	1.339	1.372	-2.5	126	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.253	18.6	142	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	0.953	-1.2	154#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	131	0.00
33	Benzo(b)fluoranthene	0.918	0.996	-8.5	150#	0.00
34	Benzo(k)fluoranthene	1.392	1.335	4.1	122	0.00
35 C	Benzo(a)pyrene	0.850	0.890	-4.7	145	0.00
36	Dibenzo(a,h)anthracene	0.766	0.806	-5.2	159#	0.00
37	Benzo(g,h,i)perylene	0.923	0.958	-3.8	140	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0