

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081015\
 Data File : BE090414.D
 Acq On : 10 Aug 2015 22:36
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 11 03:56:53 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	132	0.00
2	1,4-Dioxane	0.687	0.766	-11.5	137	0.00
3	n-Nitrosodimethylamine	0.773	0.842	-8.9	144	0.00
4 S	2-Fluorophenol	0.750	1.175	-56.7#	221#	0.00
5 S	Phenol-d6	1.241	1.746	-40.7#	200#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	138	0.00
7 S	Nitrobenzene-d5	0.331	0.375	-13.3	165#	0.00
8	Nitrobenzene	0.349	0.382	-9.5	169#	0.00
9	Naphthalene	1.104	1.127	-2.1	142	0.00
10	Hexachlorobutadiene	0.191	0.187	2.1	134	0.00
11	2-Methylnaphthalene	0.640	0.730	-14.1	165#	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	148	0.00
13 S	2,4,6-Tribromophenol	0.095	0.164	-72.6#	326#	0.00
14 S	2-Fluorobiphenyl	1.375	1.475	-7.3	156#	0.00
15	Acenaphthylene	8.098	8.880	-9.7	163#	0.00
16	Acenaphthene	1.259	1.269	-0.8	149	0.00
17	Fluorene	1.635	1.689	-3.3	151#	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	140	0.00
19	4-Bromophenyl-phenylether	0.194	0.207	-6.7	153#	0.00
20	Hexachlorobenzene	1.005	0.943	6.2	131	0.00
21	Pentachlorophenol	0.044	0.039	11.4	202#	0.00
22	Phenanthrene	1.251	1.269	-1.4	144	0.00
23	Anthracene	1.028	1.167	-13.5	164#	-0.02
24	Fluoranthene	1.231	1.410	-14.5	167#	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	141	0.00
26	Pyrene	1.084	1.252	-15.5	170#	0.00
27 S	Terphenyl-d14	0.737	0.865	-17.4	168#	0.00
28	Benzo(a)anthracene	1.108	1.202	-8.5	171#	0.00
29	Chrysene	1.339	1.373	-2.5	142	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.458	-47.3#	290#	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	1.054	-11.9	192#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	152#	0.00
33	Benzo(b)fluoranthene	0.918	1.055	-14.9	186#	0.00
34	Benzo(k)fluoranthene	1.392	1.344	3.4	143	0.00
35 C	Benzo(a)pyrene	0.850	0.963	-13.3	183#	0.00
36	Dibenzo(a,h)anthracene	0.766	0.893	-16.6	205#	0.00
37	Benzo(g,h,i)perylene	0.923	1.030	-11.6	176#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0