

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081015\
 Data File : BE090414.D
 Acq On : 10 Aug 2015 22:36
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 11 03:56:53 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	132	0.00
2	1,4-Dioxane	1.000	1.192	-19.2	137	0.00
3	n-Nitrosodimethylamine	1.000	1.089	-8.9	144	0.00
4 S	2-Fluorophenol	1.000	1.566	-56.6#	221	0.00
5 S	Phenol-d6	1.000	1.288	-28.8#	200	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	138	0.00
7 S	Nitrobenzene-d5	1.000	1.086	-8.6	165	0.00
8	Nitrobenzene	1.000	1.050	-5.0	169	0.00
9	Naphthalene	1.000	1.021	-2.1	142	0.00
10	Hexachlorobutadiene	1.000	1.059	-5.9	134	0.00
11	2-Methylnaphthalene	1.000	1.141	-14.1	165	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	148	0.00
13 S	2,4,6-Tribromophenol	1.000	1.769	-76.9#	326	0.00
14 S	2-Fluorobiphenyl	1.000	1.072	-7.2	156	0.00
15	Acenaphthylene	1.000	1.097	-9.7	163	0.00
16	Acenaphthene	1.000	1.008	-0.8	149	0.00
17	Fluorene	1.000	1.033	-3.3	151	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	140	0.00
19	4-Bromophenyl-phenylether	1.000	1.063	-6.3	153	0.00
20	Hexachlorobenzene	1.000	0.939	6.1	131	0.00
21	Pentachlorophenol	1.000	1.163	-16.3	202	0.00
22	Phenanthrene	1.000	1.014	-1.4	144	0.00
23	Anthracene	1.000	1.135	-13.5	164	-0.02
24	Fluoranthene	1.000	1.145	-14.5	167	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	141	0.00
26	Pyrene	1.000	1.155	-15.5	170	0.00
27 S	Terphenyl-d14	1.000	1.174	-17.4	168	0.00
28	Benzo(a)anthracene	1.000	1.085	-8.5	171	0.00
29	Chrysene	1.000	1.025	-2.5	142	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.611	-61.1#	290	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.080	-8.0	192	0.00
32 I	Perylene-d12	5.000	5.000	0.0	152	0.00
33	Benzo(b)fluoranthene	1.000	1.090	-9.0	186	0.00
34	Benzo(k)fluoranthene	1.000	0.965	3.5	143	0.00
35 C	Benzo(a)pyrene	1.000	1.087	-8.7	183	0.00
36	Dibenzo(a,h)anthracene	1.000	1.089	-8.9	205	0.00
37	Benzo(g,h,i)perylene	1.000	1.116	-11.6	176	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0