

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081015\
 Data File : BE090416.D
 Acq On : 11 Aug 2015 00:27
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 11 04:00:42 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	130	0.00
2	1,4-Dioxane	0.687	0.679	1.2	120	0.00
3	n-Nitrosodimethylamine	0.773	0.825	-6.7	139	0.00
4 S	2-Fluorophenol	0.750	1.175	-56.7#	218#	0.00
5 S	Phenol-d6	1.241	1.718	-38.4#	194#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	137	0.00
7 S	Nitrobenzene-d5	0.331	0.373	-12.7	162#	0.00
8	Nitrobenzene	0.349	0.381	-9.2	166#	0.00
9	Naphthalene	1.104	1.125	-1.9	140	0.00
10	Hexachlorobutadiene	0.191	0.187	2.1	133	0.00
11	2-Methylnaphthalene	0.640	0.739	-15.5	165#	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	148	0.00
13 S	2,4,6-Tribromophenol	0.095	0.166	-74.7#	331#	0.00
14 S	2-Fluorobiphenyl	1.375	1.465	-6.5	155#	0.00
15	Acenaphthylene	8.098	8.982	-10.9	165#	0.00
16	Acenaphthene	1.259	1.272	-1.0	149	0.00
17	Fluorene	1.635	1.668	-2.0	149	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	139	0.00
19	4-Bromophenyl-phenylether	0.194	0.210	-8.2	153#	0.00
20	Hexachlorobenzene	1.005	0.951	5.4	131	0.00
21	Pentachlorophenol	0.044	0.041	6.8	210#	0.00
22	Phenanthrene	1.251	1.310	-4.7	147	0.00
23	Anthracene	1.028	1.210	-17.7	168#	-0.02
24	Fluoranthene	1.231	1.438	-16.8	168#	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	141	0.00
26	Pyrene	1.084	1.265	-16.7	170#	0.00
27 S	Terphenyl-d14	0.737	0.867	-17.6	167#	0.00
28	Benzo(a)anthracene	1.108	1.218	-9.9	172#	0.00
29	Chrysene	1.339	1.387	-3.6	142	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.446	-43.4#	280#	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	1.207	-28.1#	218#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	151#	0.00
33	Benzo(b)fluoranthene	0.918	1.092	-19.0	191#	0.00
34	Benzo(k)fluoranthene	1.392	1.409	-1.2	148	0.00
35 C	Benzo(a)pyrene	0.850	1.025	-20.6#	194#	0.00
36	Dibenzo(a,h)anthracene	0.766	1.014	-32.4#	231#	0.00
37	Benzo(g,h,i)perylene	0.923	1.114	-20.7	189#	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1