

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081015\
 Data File : BE090416.D
 Acq On : 11 Aug 2015 00:27
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 11 04:00:42 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	130	0.00
2	1,4-Dioxane	1.000	1.040	-4.0	120	0.00
3	n-Nitrosodimethylamine	1.000	1.066	-6.6	139	0.00
4 S	2-Fluorophenol	1.000	1.567	-56.7#	218	0.00
5 S	Phenol-d6	1.000	1.271	-27.1#	194	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	137	0.00
7 S	Nitrobenzene-d5	1.000	1.080	-8.0	162	0.00
8	Nitrobenzene	1.000	1.047	-4.7	166	0.00
9	Naphthalene	1.000	1.019	-1.9	140	0.00
10	Hexachlorobutadiene	1.000	1.063	-6.3	133	0.00
11	2-Methylnaphthalene	1.000	1.155	-15.5	165	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	148	0.00
13 S	2,4,6-Tribromophenol	1.000	1.786	-78.6#	331	0.00
14 S	2-Fluorobiphenyl	1.000	1.065	-6.5	155	0.00
15	Acenaphthylene	1.000	1.109	-10.9	165	0.00
16	Acenaphthene	1.000	1.011	-1.1	149	0.00
17	Fluorene	1.000	1.020	-2.0	149	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	139	0.00
19	4-Bromophenyl-phenylether	1.000	1.079	-7.9	153	0.00
20	Hexachlorobenzene	1.000	0.947	5.3	131	0.00
21	Pentachlorophenol	1.000	1.208	-20.8	210	0.00
22	Phenanthrene	1.000	1.047	-4.7	147	0.00
23	Anthracene	1.000	1.177	-17.7	168	-0.02
24	Fluoranthene	1.000	1.168	-16.8	168	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	141	0.00
26	Pyrene	1.000	1.167	-16.7	170	0.00
27 S	Terphenyl-d14	1.000	1.177	-17.7	167	0.00
28	Benzo(a)anthracene	1.000	1.099	-9.9	172	0.00
29	Chrysene	1.000	1.035	-3.5	142	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.577	-57.7#	280	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.200	-20.0	218	0.00
32 I	Perylene-d12	5.000	5.000	0.0	151	0.00
33	Benzo(b)fluoranthene	1.000	1.122	-12.2	191	0.00
34	Benzo(k)fluoranthene	1.000	1.012	-1.2	148	0.00
35 C	Benzo(a)pyrene	1.000	1.142	-14.2	194	0.00
36	Dibenzo(a,h)anthracene	1.000	1.197	-19.7	231	0.00
37	Benzo(g,h,i)perylene	1.000	1.207	-20.7	189	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0