

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081015\
 Data File : BE090427.D
 Acq On : 11 Aug 2015 7:04
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 11 07:37:47 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	126	0.00
2	1,4-Dioxane	1.000	1.076	-7.6	120	0.00
3	n-Nitrosodimethylamine	1.000	1.068	-6.8	136	0.00
4 S	2-Fluorophenol	1.000	1.517	-51.7#	206	0.00
5 S	Phenol-d6	1.000	1.242	-24.2	183	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	133	0.00
7 S	Nitrobenzene-d5	1.000	1.042	-4.2	151	0.00
8	Nitrobenzene	1.000	1.017	-1.7	157	0.00
9	Naphthalene	1.000	1.027	-2.7	138	0.00
10	Hexachlorobutadiene	1.000	1.074	-7.4	131	0.00
11	2-Methylnaphthalene	1.000	1.130	-13.0	157	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	144	0.00
13 S	2,4,6-Tribromophenol	1.000	1.635	-63.5#	290	0.00
14 S	2-Fluorobiphenyl	1.000	1.055	-5.5	150	0.00
15	Acenaphthylene	1.000	1.082	-8.2	158	0.00
16	Acenaphthene	1.000	1.013	-1.3	146	0.00
17	Fluorene	1.000	1.016	-1.6	145	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	136	0.00
19	4-Bromophenyl-phenylether	1.000	1.025	-2.5	143	0.00
20	Hexachlorobenzene	1.000	0.925	7.5	125	0.00
21	Pentachlorophenol	1.000	1.193	-19.3	202	0.00
22	Phenanthrene	1.000	1.020	-2.0	140	0.00
23	Anthracene	1.000	1.121	-12.1	156	-0.02
24	Fluoranthene	1.000	1.116	-11.6	157	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	132	0.00
26	Pyrene	1.000	1.158	-15.8	159	0.00
27 S	Terphenyl-d14	1.000	1.173	-17.3	156	0.00
28	Benzo(a)anthracene	1.000	1.096	-9.6	161	0.00
29	Chrysene	1.000	1.034	-3.4	133	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.402	-40.2#	227	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.093	-9.3	181	0.00
32 I	Perylene-d12	5.000	5.000	0.0	141	0.00
33	Benzo(b)fluoranthene	1.000	1.112	-11.2	176	0.00
34	Benzo(k)fluoranthene	1.000	0.945	5.5	129	0.00
35 C	Benzo(a)pyrene	1.000	1.076	-7.6	167	0.00
36	Dibenzo(a,h)anthracene	1.000	1.092	-9.2	190	0.00
37	Benzo(g,h,i)perylene	1.000	1.112	-11.2	162	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0