

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081015\
 Data File : BE090427.D
 Acq On : 11 Aug 2015 7:04
 Operator : TP/UM
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 11 07:37:47 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	126	0.00
2	1,4-Dioxane	0.687	0.700	-1.9	120	0.00
3	n-Nitrosodimethylamine	0.773	0.826	-6.9	136	0.00
4 S	2-Fluorophenol	0.750	1.138	-51.7#	206#	0.00
5 S	Phenol-d6	1.241	1.669	-34.5#	183#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	133	0.00
7 S	Nitrobenzene-d5	0.331	0.357	-7.9	151#	0.00
8	Nitrobenzene	0.349	0.367	-5.2	157#	0.00
9	Naphthalene	1.104	1.134	-2.7	138	0.00
10	Hexachlorobutadiene	0.191	0.189	1.0	131	0.00
11	2-Methylnaphthalene	0.640	0.723	-13.0	157#	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	144	0.00
13 S	2,4,6-Tribromophenol	0.095	0.149	-56.8#	290#	0.00
14 S	2-Fluorobiphenyl	1.375	1.452	-5.6	150#	0.00
15	Acenaphthylene	8.098	8.760	-8.2	158#	0.00
16	Acenaphthene	1.259	1.275	-1.3	146	0.00
17	Fluorene	1.635	1.661	-1.6	145	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	136	0.00
19	4-Bromophenyl-phenylether	0.194	0.199	-2.6	143	0.00
20	Hexachlorobenzene	1.005	0.929	7.6	125	0.00
21	Pentachlorophenol	0.044	0.040	9.1	202#	0.00
22	Phenanthrene	1.251	1.276	-2.0	140	0.00
23	Anthracene	1.028	1.152	-12.1	156#	-0.02
24	Fluoranthene	1.231	1.374	-11.6	157#	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	132	0.00
26	Pyrene	1.084	1.255	-15.8	159#	0.00
27 S	Terphenyl-d14	0.737	0.864	-17.2	156#	0.00
28	Benzo(a)anthracene	1.108	1.214	-9.6	161#	0.00
29	Chrysene	1.339	1.386	-3.5	133	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.386	-24.1	227#	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	1.070	-13.6	181#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	141	0.00
33	Benzo(b)fluoranthene	0.918	1.081	-17.8	176#	0.00
34	Benzo(k)fluoranthene	1.392	1.316	5.5	129	0.00
35 C	Benzo(a)pyrene	0.850	0.951	-11.9	167#	0.00
36	Dibenzo(a,h)anthracene	0.766	0.897	-17.1	190#	0.00
37	Benzo(g,h,i)perylene	0.923	1.027	-11.3	162#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0