

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081016\
 Data File : BE092245.D
 Acq On : 10 Aug 2016 9:04
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 10 18:51:38 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 10 11:12:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.638	0.696	-9.1	105	0.00
3	n-Nitrosodimethylamine	0.775	0.678	12.5	101	0.00
4 S	2-Fluorophenol	1.094	0.957	12.5	99	0.00
5 S	Phenol-d6	1.573	1.478	6.0	124	0.00
6	bis(2-Chloroethyl)ether	1.275	1.126	11.7	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.353	0.348	1.4	118	-0.02
9	Naphthalene	1.121	1.105	1.4	100	0.00
10	Hexachlorobutadiene	0.162	0.168	-3.7	105	0.00
11	2-Methylnaphthalene	0.700	0.686	2.0	101	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
13 S	2,4,6-Tribromophenol	0.150	0.114	24.0	99	0.00
14 S	2-Fluorobiphenyl	1.545	1.524	1.4	101	0.00
15	Acenaphthylene	8.647	8.334	3.6	100	0.00
16	Acenaphthene	1.388	1.392	-0.3	102	0.00
17	Fluorene	1.677	1.637	2.4	100	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
19	Hexachlorobenzene	0.198	0.201	-1.5	102	0.00
20	Pentachlorophenol	0.042	0.032	23.8	94	0.02
21	Phenanthrene	1.293	1.247	3.6	101	0.00
22	Anthracene	1.169	1.050	10.2	98	0.00
23	Fluoranthene	5.137	4.791	6.7	99	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
25	Pyrene	5.914	5.617	5.0	100	0.00
26 S	Terphenyl-d14	0.845	0.796	5.8	102	0.02
27	Benzo(a)anthracene	6.002	5.799	3.4	104	0.00
28	Chrysene	5.546	6.169	-11.2	110	0.00
29	Bis(2-ethylhexyl)phthalate	0.770	0.583	24.3	93	0.00
30	Indeno(1,2,3-cd)pyrene	5.972	5.352	10.4	99	0.00
31 I	Perylene-d12	1.000	1.000	0.0	101	0.00
32	Benzo(b)fluoranthene	1.313	1.324	-0.8	111	0.00
33	Benzo(k)fluoranthene	1.323	1.170	11.6	89	0.01
34 C	Benzo(a)pyrene	1.238	1.116	9.9	100	0.00
35	Dibenzo(a,h)anthracene	1.151	1.018	11.6	100	0.00
36	Benzo(g,h,i)perylene	4.954	4.580	7.5	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0