

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081115\
 Data File : BE090429.D
 Acq On : 11 Aug 2015 15:40
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 12 04:18:51 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.687	0.709	-3.2	116	0.00
3	n-Nitrosodimethylamine	0.773	0.802	-3.8	125	0.00
4 S	2-Fluorophenol	0.750	1.046	-39.5#	180#	0.00
5 S	Phenol-d6	1.241	1.620	-30.5#	169#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
7 S	Nitrobenzene-d5	0.331	0.349	-5.4	142	0.00
8	Nitrobenzene	0.349	0.356	-2.0	145	0.00
9	Naphthalene	1.104	1.118	-1.3	130	0.00
10	Hexachlorobutadiene	0.191	0.189	1.0	125	0.00
11	2-Methylnaphthalene	0.640	0.716	-11.9	149	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	140	0.00
13 S	2,4,6-Tribromophenol	0.095	0.130	-36.8#	245#	0.00
14 S	2-Fluorobiphenyl	1.375	1.437	-4.5	144	0.00
15	Acenaphthylene	8.098	8.556	-5.7	149	0.00
16	Acenaphthene	1.259	1.260	-0.1	140	0.00
17	Fluorene	1.635	1.665	-1.8	141	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	134	0.00
19	4-Bromophenyl-phenylether	0.194	0.195	-0.5	137	0.00
20	Hexachlorobenzene	1.005	0.936	6.9	125	0.00
21	Pentachlorophenol	0.044	0.036	18.2	176#	0.00
22	Phenanthrene	1.251	1.246	0.4	135	0.00
23	Anthracene	1.028	1.126	-9.5	151#	-0.02
24	Fluoranthene	1.231	1.324	-7.6	150	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	131	0.00
26	Pyrene	1.084	1.200	-10.7	150#	0.00
27 S	Terphenyl-d14	0.737	0.828	-12.3	148	0.00
28	Benzo(a)anthracene	1.108	1.169	-5.5	153#	0.00
29	Chrysene	1.339	1.366	-2.0	130	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.311	0.0	182#	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	1.021	-8.4	172#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	136	0.00
33	Benzo(b)fluoranthene	0.918	1.020	-11.1	160#	0.00
34	Benzo(k)fluoranthene	1.392	1.349	3.1	128	0.00
35 C	Benzo(a)pyrene	0.850	0.919	-8.1	156#	0.00
36	Dibenzo(a,h)anthracene	0.766	0.869	-13.4	178#	0.00
37	Benzo(g,h,i)perylene	0.923	0.991	-7.4	151#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0