

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081115\
 Data File : BE090429.D
 Acq On : 11 Aug 2015 15:40
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 12 04:18:51 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	120	0.00
2	1,4-Dioxane	1.000	1.092	-9.2	116	0.00
3	n-Nitrosodimethylamine	1.000	1.037	-3.7	125	0.00
4 S	2-Fluorophenol	1.000	1.395	-39.5#	180	0.00
5 S	Phenol-d6	1.000	1.213	-21.3	169	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	128	0.00
7 S	Nitrobenzene-d5	1.000	1.024	-2.4	142	0.00
8	Nitrobenzene	1.000	0.991	0.9	145	0.00
9	Naphthalene	1.000	1.013	-1.3	130	0.00
10	Hexachlorobutadiene	1.000	1.074	-7.4	125	0.00
11	2-Methylnaphthalene	1.000	1.118	-11.8	149	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	140	0.00
13 S	2,4,6-Tribromophenol	1.000	1.464	-46.4#	245	0.00
14 S	2-Fluorobiphenyl	1.000	1.045	-4.5	144	0.00
15	Acenaphthylene	1.000	1.057	-5.7	149	0.00
16	Acenaphthene	1.000	1.001	-0.1	140	0.00
17	Fluorene	1.000	1.018	-1.8	141	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	134	0.00
19	4-Bromophenyl-phenylether	1.000	1.001	-0.1	137	0.00
20	Hexachlorobenzene	1.000	0.932	6.8	125	0.00
21	Pentachlorophenol	1.000	1.082	-8.2	176	0.00
22	Phenanthrene	1.000	0.996	0.4	135	0.00
23	Anthracene	1.000	1.096	-9.6	151	-0.02
24	Fluoranthene	1.000	1.075	-7.5	150	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	131	0.00
26	Pyrene	1.000	1.108	-10.8	150	0.00
27 S	Terphenyl-d14	1.000	1.125	-12.5	148	0.00
28	Benzo(a)anthracene	1.000	1.055	-5.5	153	0.00
29	Chrysene	1.000	1.020	-2.0	130	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.177	-17.7	182	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.054	-5.4	172	0.00
32 I	Perylene-d12	5.000	5.000	0.0	136	0.00
33	Benzo(b)fluoranthene	1.000	1.061	-6.1	160	0.00
34	Benzo(k)fluoranthene	1.000	0.969	3.1	128	0.00
35 C	Benzo(a)pyrene	1.000	1.047	-4.7	156	0.00
36	Dibenzo(a,h)anthracene	1.000	1.067	-6.7	178	0.00
37	Benzo(g,h,i)perylene	1.000	1.073	-7.3	151	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0