

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081115\
 Data File : BE090443.D
 Acq On : 12 Aug 2015 00:13
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 12 05:46:15 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	120	0.00
2	1,4-Dioxane	1.000	0.973	2.7	104	0.00
3	n-Nitrosodimethylamine	1.000	1.015	-1.5	122	0.00
4 S	2-Fluorophenol	1.000	1.402	-40.2#	180	0.00
5 S	Phenol-d6	1.000	1.204	-20.4	167	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	126	0.00
7 S	Nitrobenzene-d5	1.000	1.022	-2.2	139	0.00
8	Nitrobenzene	1.000	1.011	-1.1	146	0.00
9	Naphthalene	1.000	1.018	-1.8	129	0.00
10	Hexachlorobutadiene	1.000	1.091	-9.1	125	0.00
11	2-Methylnaphthalene	1.000	1.095	-9.5	144	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	134	0.00
13 S	2,4,6-Tribromophenol	1.000	1.465	-46.5#	236	0.00
14 S	2-Fluorobiphenyl	1.000	1.050	-5.0	139	0.00
15	Acenaphthylene	1.000	1.052	-5.2	142	0.00
16	Acenaphthene	1.000	1.013	-1.3	136	0.00
17	Fluorene	1.000	1.029	-2.9	137	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	129	0.00
19	4-Bromophenyl-phenylether	1.000	1.030	-3.0	136	0.00
20	Hexachlorobenzene	1.000	0.938	6.2	120	0.00
21	Pentachlorophenol	1.000	1.236	-23.6	201	0.00
22	Phenanthrene	1.000	1.032	-3.2	134	0.00
23	Anthracene	1.000	1.116	-11.6	147	-0.02
24	Fluoranthene	1.000	1.069	-6.9	143	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	127	0.00
26	Pyrene	1.000	1.098	-9.8	145	0.00
27 S	Terphenyl-d14	1.000	1.118	-11.8	144	-0.01
28	Benzo(a)anthracene	1.000	1.060	-6.0	150	0.00
29	Chrysene	1.000	1.006	-0.6	125	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.210	-21.0	184	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.049	-4.9	166	0.00
32 I	Perylene-d12	5.000	5.000	0.0	132	0.00
33	Benzo(b)fluoranthene	1.000	1.086	-8.6	160	0.00
34	Benzo(k)fluoranthene	1.000	0.946	5.4	121	0.00
35 C	Benzo(a)pyrene	1.000	1.053	-5.3	153	0.00
36	Dibenzo(a,h)anthracene	1.000	1.066	-6.6	173	0.00
37	Benzo(g,h,i)perylene	1.000	1.063	-6.3	145	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0