

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081115\
 Data File : BE090445.D
 Acq On : 12 Aug 2015 1:37
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 12 06:42:22 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	-0.01
2	1,4-Dioxane	0.687	0.702	-2.2	111	0.00
3	n-Nitrosodimethylamine	0.773	0.809	-4.7	123	0.00
4 S	2-Fluorophenol	0.750	1.022	-36.3#	171#	0.00
5 S	Phenol-d6	1.241	1.585	-27.7#	161#	-0.01
6 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
7 S	Nitrobenzene-d5	0.331	0.344	-3.9	135	-0.01
8	Nitrobenzene	0.349	0.359	-2.9	142	0.00
9	Naphthalene	1.104	1.116	-1.1	126	0.00
10	Hexachlorobutadiene	0.191	0.192	-0.5	123	0.00
11	2-Methylnaphthalene	0.640	0.695	-8.6	140	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	133	0.00
13 S	2,4,6-Tribromophenol	0.095	0.123	-29.5#	220#	0.00
14 S	2-Fluorobiphenyl	1.375	1.420	-3.3	135	0.00
15	Acenaphthylene	8.098	8.438	-4.2	140	0.00
16	Acenaphthene	1.259	1.271	-1.0	134	0.00
17	Fluorene	1.635	1.669	-2.1	134	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
19	4-Bromophenyl-phenylether	0.194	0.192	1.0	132	0.00
20	Hexachlorobenzene	1.005	0.946	5.9	123	0.00
21	Pentachlorophenol	0.044	0.041	6.8	199#	0.00
22	Phenanthrene	1.251	1.238	1.0	131	0.00
23	Anthracene	1.028	1.107	-7.7	145	-0.02
24	Fluoranthene	1.231	1.273	-3.4	141	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	127	0.00
26	Pyrene	1.084	1.189	-9.7	144	0.00
27 S	Terphenyl-d14	0.737	0.810	-9.9	140	0.00
28	Benzo(a)anthracene	1.108	1.160	-4.7	147	0.00
29	Chrysene	1.339	1.358	-1.4	125	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.289	7.1	164#	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	1.006	-6.8	164#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	130	0.00
33	Benzo(b)fluoranthene	0.918	1.026	-11.8	154#	0.00
34	Benzo(k)fluoranthene	1.392	1.327	4.7	120	0.00
35 C	Benzo(a)pyrene	0.850	0.906	-6.6	147	0.00
36	Dibenzo(a,h)anthracene	0.766	0.863	-12.7	169#	0.00
37	Benzo(g,h,i)perylene	0.923	0.979	-6.1	143	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0