

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081115\
 Data File : BE090460.D
 Acq On : 12 Aug 2015 10:37
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 12 11:25:24 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2	1,4-Dioxane	0.687	0.674	1.9	110	0.00
3	n-Nitrosodimethylamine	0.773	0.828	-7.1	130	0.00
4 S	2-Fluorophenol	0.750	1.072	-42.9#	185#	0.00
5 S	Phenol-d6	1.241	1.668	-34.4#	175#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
7 S	Nitrobenzene-d5	0.331	0.359	-8.5	146	0.00
8	Nitrobenzene	0.349	0.371	-6.3	151#	0.00
9	Naphthalene	1.104	1.128	-2.2	131	0.00
10	Hexachlorobutadiene	0.191	0.192	-0.5	127	0.00
11	2-Methylnaphthalene	0.640	0.693	-8.3	144	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	136	0.00
13 S	2,4,6-Tribromophenol	0.095	0.133	-40.0#	244#	0.00
14 S	2-Fluorobiphenyl	1.375	1.423	-3.5	139	0.00
15	Acenaphthylene	8.098	8.406	-3.8	143	0.00
16	Acenaphthene	1.259	1.266	-0.6	137	0.00
17	Fluorene	1.635	1.648	-0.8	136	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	132	0.00
19	4-Bromophenyl-phenylether	0.194	0.195	-0.5	136	0.00
20	Hexachlorobenzene	1.005	0.931	7.4	122	0.00
21	Pentachlorophenol	0.044	0.045	-2.3	219#	0.00
22	Phenanthrene	1.251	1.259	-0.6	134	0.00
23	Anthracene	1.028	1.134	-10.3	150	-0.02
24	Fluoranthene	1.231	1.301	-5.7	145	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	129	0.00
26	Pyrene	1.084	1.197	-10.4	148	0.00
27 S	Terphenyl-d14	0.737	0.817	-10.9	145	-0.01
28	Benzo(a)anthracene	1.108	1.188	-7.2	154#	0.00
29	Chrysene	1.339	1.316	1.7	124	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.322	-3.5	186#	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	1.016	-7.9	169#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	135	0.00
33	Benzo(b)fluoranthene	0.918	1.047	-14.1	163#	0.00
34	Benzo(k)fluoranthene	1.392	1.248	10.3	117	0.00
35 C	Benzo(a)pyrene	0.850	0.925	-8.8	156#	0.00
36	Dibenzo(a,h)anthracene	0.766	0.867	-13.2	176#	-0.01
37	Benzo(g,h,i)perylene	0.923	0.970	-5.1	146	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0