

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081116\
 Data File : BE092263.D
 Acq On : 11 Aug 2016 12:46
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 11 17:12:21 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 10 11:12:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	147	0.00
2	1,4-Dioxane	0.638	0.627	1.7	135	0.00
3	n-Nitrosodimethylamine	0.775	0.752	3.0	160#	0.00
4 S	2-Fluorophenol	1.094	1.085	0.8	160#	0.00
5 S	Phenol-d6	1.573	1.456	7.4	174#	-0.02
6	bis(2-Chloroethyl)ether	1.275	1.340	-5.1	171#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	145	0.00
8 S	Nitrobenzene-d5	0.353	0.325	7.9	159#	-0.02
9	Naphthalene	1.121	1.117	0.4	145	-0.02
10	Hexachlorobutadiene	0.162	0.166	-2.5	150	-0.02
11	2-Methylnaphthalene	0.700	0.700	0.0	148	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	140	0.00
13 S	2,4,6-Tribromophenol	0.150	0.128	14.7	155#	0.00
14 S	2-Fluorobiphenyl	1.545	1.551	-0.4	145	-0.01
15	Acenaphthylene	8.647	8.505	1.6	144	0.00
16	Acenaphthene	1.388	1.379	0.6	142	0.00
17	Fluorene	1.677	1.674	0.2	143	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	142	0.00
19	Hexachlorobenzene	0.198	0.186	6.1	133	0.00
20	Pentachlorophenol	0.042	0.047	-11.9	197#	0.00
21	Phenanthrene	1.293	1.273	1.5	145	0.00
22	Anthracene	1.169	1.084	7.3	142	0.00
23	Fluoranthene	5.137	4.868	5.2	142	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	149	-0.02
25	Pyrene	5.914	5.724	3.2	151#	0.00
26 S	Terphenyl-d14	0.845	0.819	3.1	156#	0.00
27	Benzo(a)anthracene	6.002	6.145	-2.4	164#	-0.02
28	Chrysene	5.546	5.246	5.4	139	0.00
29	Bis(2-ethylhexyl)phthalate	0.770	0.596	22.6	141	0.00
30	Indeno(1,2,3-cd)pyrene	5.972	5.785	3.1	159#	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	149	0.00
32	Benzo(b)fluoranthene	1.313	1.301	0.9	160#	0.00
33	Benzo(k)fluoranthene	1.323	1.293	2.3	145	0.00
34 C	Benzo(a)pyrene	1.238	1.227	0.9	161#	0.00
35	Dibenzo(a,h)anthracene	1.151	1.130	1.8	163#	-0.02
36	Benzo(g,h,i)perylene	4.954	4.754	4.0	151#	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0