

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081116\
 Data File : BE092267.D
 Acq On : 11 Aug 2016 15:15
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 11 17:11:48 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 10 11:12:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	146	0.00
2	1,4-Dioxane	0.638	0.636	0.3	136	0.00
3	n-Nitrosodimethylamine	0.775	0.760	1.9	161#	-0.01
4 S	2-Fluorophenol	1.094	1.088	0.5	160#	0.00
5 S	Phenol-d6	1.573	1.498	4.8	178#	-0.02
6	bis(2-Chloroethyl)ether	1.275	1.374	-7.8	175#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	145	0.00
8 S	Nitrobenzene-d5	0.353	0.332	5.9	161#	-0.02
9	Naphthalene	1.121	1.115	0.5	145	-0.02
10	Hexachlorobutadiene	0.162	0.163	-0.6	147	-0.02
11	2-Methylnaphthalene	0.700	0.701	-0.1	148	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	143	0.00
13 S	2,4,6-Tribromophenol	0.150	0.128	14.7	158#	0.00
14 S	2-Fluorobiphenyl	1.545	1.522	1.5	145	-0.01
15	Acenaphthylene	8.647	8.399	2.9	145	0.00
16	Acenaphthene	1.388	1.377	0.8	144	0.00
17	Fluorene	1.677	1.652	1.5	144	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	145	0.00
19	Hexachlorobenzene	0.198	0.180	9.1	131	0.00
20	Pentachlorophenol	0.042	0.044	-4.8	186#	0.00
21	Phenanthrene	1.293	1.208	6.6	140	0.00
22	Anthracene	1.169	1.099	6.0	147	0.00
23	Fluoranthene	5.137	4.807	6.4	143	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	147	-0.02
25	Pyrene	5.914	6.028	-1.9	157#	0.00
26 S	Terphenyl-d14	0.845	0.858	-1.5	161#	0.00
27	Benzo(a)anthracene	6.002	6.504	-8.4	171#	-0.02
28	Chrysene	5.546	5.452	1.7	142	0.00
29	Bis(2-ethylhexyl)phthalate	0.770	0.670	13.0	157#	0.00
30	Indeno(1,2,3-cd)pyrene	5.972	6.059	-1.5	164#	-0.02
31 I	Perylene-d12	1.000	1.000	0.0	148	0.00
32	Benzo(b)fluoranthene	1.313	1.308	0.4	160#	0.00
33	Benzo(k)fluoranthene	1.323	1.346	-1.7	150	0.00
34 C	Benzo(a)pyrene	1.238	1.265	-2.2	166#	0.00
35	Dibenzo(a,h)anthracene	1.151	1.163	-1.0	167#	0.00
36	Benzo(g,h,i)perylene	4.954	4.852	2.1	153#	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0