

Data Path : Z:\HPCHEM1\BNA_E\Data\BE081117\
 Data File : BE093771.D
 Acq On : 11 Aug 2017 14:09
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.401

Quant Time: Aug 12 01:50:59 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 11 18:19:02 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	81	0.00
3	Naphthalene	0.973	0.991	-1.8	84	0.00
4 SURR	2-Methylnaphthalene-d10	0.648	0.644	0.6	82	0.00
5	2-Methylnaphthalene	0.710	0.700	1.4	81	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	72	0.00
7	Acenaphthylene	1.673	1.701	-1.7	76	0.00
8 C	Acenaphthene	1.316	1.305	0.8	73	0.00
9	Fluorene	1.664	1.494	10.2	67	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	60	0.00
11	Pentachlorophenol	0.066	0.055	16.7	56	0.00
12	Phenanthrene	1.089	1.069	1.8	61	0.00
13	Anthracene	1.061	1.011	4.7	60	0.00
14 SURR	Fluoranthene-d10	1.263	1.101	12.8	54	0.00
15 C	Fluoranthene	1.362	1.182	13.2	55	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	54	0.00
17	Pyrene	1.168	1.171	-0.3	55	0.00
18	Benzo(a)anthracene	1.160	1.121	3.4	54	0.00
19	Chrysene	1.112	1.114	-0.2	55	0.00
20 I	Perylene-d12	1.000	1.000	0.0	55	0.00
21	Benzo(b)fluoranthene	1.237	1.232	0.4	57	0.00
22	Benzo(k)fluoranthene	1.220	1.152	5.6	54	0.00
23 C	Benzo(a)pyrene	1.177	1.170	0.6	57	0.00
24	Indeno(1,2,3-cd)pyrene	1.299	1.258	3.2	55	0.00
25	Dibenzo(a,h)anthracene	1.086	1.049	3.4	56	0.00
26	Benzo(g,h,i)perylene	1.095	1.024	6.5	53	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0