

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081215\
 Data File : BE090464.D
 Acq On : 12 Aug 2015 13:06
 Operator : UM/IZ
 Sample : G3236-13
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 080915-D534-E-D-B

Manual Integrations
 APPROVED

MMDadoda
 8/13/2015 3:35:11 PM

Quant Time: Aug 13 12:49:56 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	20279	5.00	ng	0.00
6) Naphthalene-d8	10.34	136	93114	5.00	ng	0.00
12) Acenaphthene-d10	14.19	164	53917	5.00	ng	0.00
18) Phenanthrene-d10	16.94	188	128007	5.00	ng	0.00
25) Chrysene-d12	21.09	240	149038	5.00	ng	0.00
32) Perylene-d12	23.39	264	132105	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.21	112	14323	4.71	ng	0.00
5) Phenol-d6	6.77	99	12911	2.14	ng	0.00
7) Nitrobenzene-d5	8.73	82	26353	3.54	ng	0.00
13) 2,4,6-Tribromophenol	15.71	330	16081	9.13	ng	0.00
14) 2-Fluorobiphenyl	12.82	172	64733	4.37	ng	0.00
27) Terphenyl-d14	19.55	244	125804	5.73	ng	-0.01
Target Compounds						
9) Naphthalene	10.38	128	520	0.03	ng	Qvalue # 62
22) Phenanthrene	16.98	178	1220m	0.04	ng	
26) Pyrene	19.34	202	1008	0.03	ng	96
28) Benzo(a)anthracene	21.09	228	1100	0.03	ng	# 85
29) Chrysene	21.13	228	889	0.02	ng	# 82
30) Bis(2-ethylhexyl)phthalate	21.02	149	2435	0.41	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081215\
Data File : BE090464.D
Acq On : 12 Aug 2015 13:06
Operator : UM/IZ
Sample : G3236-13
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
080915-D534-E-D-B

Manual Integrations
APPROVED

MMDadoda
8/13/2015 3:35:11 PM

Quant Time: Aug 13 12:49:56 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 06 12:28:08 2015
Response via : Initial Calibration

