

Data Path : Z:\HPCHEM1\BNA_E\Data\BE081215\
 Data File : BE090462.D
 Acq On : 12 Aug 2015 11:51
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 13 01:22:15 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	118	0.00
2	1,4-Dioxane	1.000	1.002	-0.2	105	0.00
3	n-Nitrosodimethylamine	1.000	1.041	-4.1	123	0.00
4 S	2-Fluorophenol	1.000	1.367	-36.7#	173	0.00
5 S	Phenol-d6	1.000	1.199	-19.9	164	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	125	0.00
7 S	Nitrobenzene-d5	1.000	1.036	-3.6	140	0.00
8	Nitrobenzene	1.000	1.007	-0.7	144	0.00
9	Naphthalene	1.000	1.019	-1.9	128	0.00
10	Hexachlorobutadiene	1.000	1.086	-8.6	123	0.00
11	2-Methylnaphthalene	1.000	1.058	-5.8	138	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	134	0.00
13 S	2,4,6-Tribromophenol	1.000	1.346	-34.6#	212	0.00
14 S	2-Fluorobiphenyl	1.000	1.023	-2.3	134	0.00
15	Acenaphthylene	1.000	1.013	-1.3	136	0.00
16	Acenaphthene	1.000	0.990	1.0	132	0.00
17	Fluorene	1.000	1.010	-1.0	133	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	128	0.00
19	4-Bromophenyl-phenylether	1.000	1.007	-0.7	132	0.00
20	Hexachlorobenzene	1.000	0.929	7.1	118	0.00
21	Pentachlorophenol	1.000	1.212	-21.2	195	0.00
22	Phenanthrene	1.000	1.012	-1.2	131	0.00
23	Anthracene	1.000	1.075	-7.5	141	-0.02
24	Fluoranthene	1.000	1.036	-3.6	138	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	126	0.00
26	Pyrene	1.000	1.075	-7.5	140	0.00
27 S	Terphenyl-d14	1.000	1.076	-7.6	136	-0.01
28	Benzo(a)anthracene	1.000	1.042	-4.2	146	0.00
29	Chrysene	1.000	0.972	2.8	119	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.077	-7.7	157	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.036	-3.6	161	0.00
32 I	Perylene-d12	5.000	5.000	0.0	129	0.00
33	Benzo(b)fluoranthene	1.000	1.067	-6.7	153	0.00
34	Benzo(k)fluoranthene	1.000	0.901	9.9	113	0.00
35 C	Benzo(a)pyrene	1.000	1.035	-3.5	146	0.00
36	Dibenzo(a,h)anthracene	1.000	1.061	-6.1	168	0.00
37	Benzo(g,h,i)perylene	1.000	1.037	-3.7	138	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0