

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081215\
 Data File : BE090462.D
 Acq On : 12 Aug 2015 11:51
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 13 01:22:15 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	0.00
2	1,4-Dioxane	0.687	0.657	4.4	105	0.00
3	n-Nitrosodimethylamine	0.773	0.805	-4.1	123	0.00
4 S	2-Fluorophenol	0.750	1.025	-36.7#	173#	0.00
5 S	Phenol-d6	1.241	1.597	-28.7#	164#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	125	0.00
7 S	Nitrobenzene-d5	0.331	0.354	-6.9	140	0.00
8	Nitrobenzene	0.349	0.363	-4.0	144	0.00
9	Naphthalene	1.104	1.125	-1.9	128	0.00
10	Hexachlorobutadiene	0.191	0.191	0.0	123	0.00
11	2-Methylnaphthalene	0.640	0.677	-5.8	138	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	134	0.00
13 S	2,4,6-Tribromophenol	0.095	0.118	-24.2	212#	0.00
14 S	2-Fluorobiphenyl	1.375	1.407	-2.3	134	0.00
15	Acenaphthylene	8.098	8.202	-1.3	136	0.00
16	Acenaphthene	1.259	1.246	1.0	132	0.00
17	Fluorene	1.635	1.652	-1.0	133	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	128	0.00
19	4-Bromophenyl-phenylether	0.194	0.196	-1.0	132	0.00
20	Hexachlorobenzene	1.005	0.933	7.2	118	0.00
21	Pentachlorophenol	0.044	0.041	6.8	195#	0.00
22	Phenanthrene	1.251	1.266	-1.2	131	0.00
23	Anthracene	1.028	1.106	-7.6	141	-0.02
24	Fluoranthene	1.231	1.275	-3.6	138	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	126	0.00
26	Pyrene	1.084	1.165	-7.5	140	0.00
27 S	Terphenyl-d14	0.737	0.792	-7.5	136	-0.01
28	Benzo(a)anthracene	1.108	1.155	-4.2	146	0.00
29	Chrysene	1.339	1.303	2.7	119	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.279	10.3	157#	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	0.998	-5.9	161#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	129	0.00
33	Benzo(b)fluoranthene	0.918	1.028	-12.0	153#	0.00
34	Benzo(k)fluoranthene	1.392	1.254	9.9	113	0.00
35 C	Benzo(a)pyrene	0.850	0.906	-6.6	146	0.00
36	Dibenzo(a,h)anthracene	0.766	0.862	-12.5	168#	0.00
37	Benzo(g,h,i)perylene	0.923	0.957	-3.7	138	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0