

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081215\
 Data File : BE090478.D
 Acq On : 12 Aug 2015 21:45
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 13 01:22:01 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
2	1,4-Dioxane	0.687	0.698	-1.6	103	0.00
3	n-Nitrosodimethylamine	0.773	0.811	-4.9	114	0.00
4 S	2-Fluorophenol	0.750	1.079	-43.9#	167#	0.00
5 S	Phenol-d6	1.241	1.537	-23.9	145	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	109	0.00
7 S	Nitrobenzene-d5	0.331	0.356	-7.6	123	0.00
8	Nitrobenzene	0.349	0.372	-6.6	129	0.00
9	Naphthalene	1.104	1.124	-1.8	112	0.00
10	Hexachlorobutadiene	0.191	0.192	-0.5	109	0.00
11	2-Methylnaphthalene	0.640	0.695	-8.6	123	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
13 S	2,4,6-Tribromophenol	0.095	0.140	-47.4#	224#	0.00
14 S	2-Fluorobiphenyl	1.375	1.425	-3.6	121	0.00
15	Acenaphthylene	8.098	8.432	-4.1	125	0.00
16	Acenaphthene	1.259	1.270	-0.9	120	0.00
17	Fluorene	1.635	1.685	-3.1	121	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	119	0.00
19	4-Bromophenyl-phenylether	0.194	0.192	1.0	120	0.00
20	Hexachlorobenzene	1.005	0.943	6.2	111	0.00
21	Pentachlorophenol	0.044	0.051	-15.9	225#	0.00
22	Phenanthrene	1.251	1.276	-2.0	122	0.00
23	Anthracene	1.028	1.130	-9.9	134	-0.02
24	Fluoranthene	1.231	1.360	-10.5	136	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	125	0.00
26	Pyrene	1.084	1.155	-6.5	138	0.00
27 S	Terphenyl-d14	0.737	0.802	-8.8	137	-0.01
28	Benzo(a)anthracene	1.108	1.200	-8.3	150#	0.00
29	Chrysene	1.339	1.308	2.3	119	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.345	-10.9	192#	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	1.098	-16.6	176#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	135	0.00
33	Benzo(b)fluoranthene	0.918	1.066	-16.1	166#	0.00
34	Benzo(k)fluoranthene	1.392	1.251	10.1	117	0.00
35 C	Benzo(a)pyrene	0.850	0.948	-11.5	160#	0.00
36	Dibenzo(a,h)anthracene	0.766	0.896	-17.0	182#	-0.01
37	Benzo(g,h,i)perylene	0.923	1.007	-9.1	152#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0