

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081215\
 Data File : BE090497.D
 Acq On : 13 Aug 2015 9:11
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 13 10:51:00 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area	% Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	137	0.00
2	1,4-Dioxane	1.000	1.033	-3.3	125	0.00
3	n-Nitrosodimethylamine	1.000	1.104	-10.4	152	0.00
4 S	2-Fluorophenol	1.000	2.003	-100.3#	294	0.00
5 S	Phenol-d6	1.000	1.589	-58.9#	268	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	143	0.00
7 S	Nitrobenzene-d5	1.000	1.222	-22.2	197	-0.01
8	Nitrobenzene	1.000	1.202	-20.2	206	0.00
9	Naphthalene	1.000	1.014	-1.4	146	0.00
10	Hexachlorobutadiene	1.000	1.039	-3.9	136	0.00
11	2-Methylnaphthalene	1.000	1.186	-18.6	177	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	151	0.00
13 S	2,4,6-Tribromophenol	1.000	2.464	-146.4#	502	-0.02
14 S	2-Fluorobiphenyl	1.000	1.084	-8.4	161	0.00
15	Acenaphthylene	1.000	1.141	-14.1	173	0.00
16	Acenaphthene	1.000	1.029	-2.9	155	0.00
17	Fluorene	1.000	1.063	-6.3	158	-0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	140	0.00
19	4-Bromophenyl-phenylether	1.000	1.127	-12.7	161	0.00
20	Hexachlorobenzene	1.000	0.921	7.9	128	-0.02
21	Pentachlorophenol	1.000	2.591	-159.1#	589	-0.01
22	Phenanthrene	1.000	1.045	-4.5	147	0.00
23	Anthracene	1.000	1.225	-22.5	176	-0.02
24	Fluoranthene	1.000	1.214	-21.4	176	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	137	0.00
26	Pyrene	1.000	1.261	-26.1#	179	0.00
27 S	Terphenyl-d14	1.000	1.302	-30.2#	180	-0.01
28	Benzo(a)anthracene	1.000	1.200	-20.0	183	0.00
29	Chrysene	1.000	1.012	-1.2	135	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	2.733	-173.3#	545	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.305	-30.5#	235	0.00
32 I	Perylene-d12	5.000	5.000	0.0	159	0.00
33	Benzo(b)fluoranthene	1.000	1.162	-16.2	210	0.00
34	Benzo(k)fluoranthene	1.000	0.958	4.2	148	0.00
35 C	Benzo(a)pyrene	1.000	1.215	-21.5#	221	0.00
36	Dibenzo(a,h)anthracene	1.000	1.243	-24.3	256	-0.01
37	Benzo(g,h,i)perylene	1.000	1.254	-25.4#	207	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1