

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081215\
 Data File : BE090497.D
 Acq On : 13 Aug 2015 9:11
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 13 10:51:00 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	137	0.00
2	1,4-Dioxane	0.687	0.675	1.7	125	0.00
3	n-Nitrosodimethylamine	0.773	0.854	-10.5	152#	0.00
4 S	2-Fluorophenol	0.750	1.502	-100.3#	294#	0.00
5 S	Phenol-d6	1.241	2.251	-81.4#	268#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	143	0.00
7 S	Nitrobenzene-d5	0.331	0.433	-30.8#	197#	-0.01
8	Nitrobenzene	0.349	0.451	-29.2#	206#	0.00
9	Naphthalene	1.104	1.119	-1.4	146	0.00
10	Hexachlorobutadiene	0.191	0.183	4.2	136	0.00
11	2-Methylnaphthalene	0.640	0.759	-18.6	177#	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	151#	0.00
13 S	2,4,6-Tribromophenol	0.095	0.247	-160.0#	502#	-0.02
14 S	2-Fluorobiphenyl	1.375	1.490	-8.4	161#	0.00
15	Acenaphthylene	8.098	9.240	-14.1	173#	0.00
16	Acenaphthene	1.259	1.295	-2.9	155#	0.00
17	Fluorene	1.635	1.737	-6.2	158#	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	140	0.00
19	4-Bromophenyl-phenylether	0.194	0.219	-12.9	161#	0.00
20	Hexachlorobenzene	1.005	0.925	8.0	128	-0.02
21	Pentachlorophenol	0.044	0.114	-159.1#	589#	-0.01
22	Phenanthrene	1.251	1.308	-4.6	147	0.00
23	Anthracene	1.028	1.259	-22.5	176#	-0.02
24	Fluoranthene	1.231	1.495	-21.4	176#	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	137	0.00
26	Pyrene	1.084	1.366	-26.0#	179#	0.00
27 S	Terphenyl-d14	0.737	0.959	-30.1#	180#	-0.01
28	Benzo(a)anthracene	1.108	1.330	-20.0	183#	0.00
29	Chrysene	1.339	1.356	-1.3	135	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.891	-186.5#	545#	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	1.340	-42.3#	235#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	159#	0.00
33	Benzo(b)fluoranthene	0.918	1.140	-24.2	210#	0.00
34	Benzo(k)fluoranthene	1.392	1.334	4.2	148	0.00
35 C	Benzo(a)pyrene	0.850	1.107	-30.2#	221#	0.00
36	Dibenzo(a,h)anthracene	0.766	1.065	-39.0#	256#	-0.01
37	Benzo(g,h,i)perylene	0.923	1.157	-25.4#	207#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1