

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081315\
 Data File : BE090499.D
 Acq On : 13 Aug 2015 13:37
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 13 15:11:26 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	130	-0.01
2	1,4-Dioxane	0.687	0.709	-3.2	125	0.00
3	n-Nitrosodimethylamine	0.773	0.865	-11.9	146	0.00
4 S	2-Fluorophenol	0.750	1.496	-99.5#	278#	0.00
5 S	Phenol-d6	1.241	2.205	-77.7#	249#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	145	0.00
7 S	Nitrobenzene-d5	0.331	0.403	-21.8	186#	-0.01
8	Nitrobenzene	0.349	0.422	-20.9	196#	-0.01
9	Naphthalene	1.104	1.132	-2.5	150	0.00
10	Hexachlorobutadiene	0.191	0.166	13.1	125	0.00
11	2-Methylnaphthalene	0.640	0.742	-15.9	176#	-0.01
12 I	Acenaphthene-d10	1.000	1.000	0.0	144	0.00
13 S	2,4,6-Tribromophenol	0.095	0.207	-117.9#	402#	-0.02
14 S	2-Fluorobiphenyl	1.375	1.503	-9.3	155#	0.00
15	Acenaphthylene	8.098	9.116	-12.6	163#	0.00
16	Acenaphthene	1.259	1.372	-9.0	156#	0.00
17	Fluorene	1.635	1.683	-2.9	146	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	134	0.00
19	4-Bromophenyl-phenylether	0.194	0.206	-6.2	146	0.00
20	Hexachlorobenzene	1.005	0.904	10.0	120	0.00
21	Pentachlorophenol	0.044	0.081	-84.1#	401#	0.00
22	Phenanthrene	1.251	1.303	-4.2	141	0.00
23	Anthracene	1.028	1.209	-17.6	162#	-0.02
24	Fluoranthene	1.231	1.420	-15.4	161#	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
26	Pyrene	1.084	1.449	-33.7#	164#	0.00
27 S	Terphenyl-d14	0.737	0.925	-25.5#	150	0.00
28	Benzo(a)anthracene	1.108	1.297	-17.1	154#	0.00
29	Chrysene	1.339	1.337	0.1	115	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.746	-139.9#	395#	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	1.282	-36.1#	195#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	128	0.00
33	Benzo(b)fluoranthene	0.918	1.202	-30.9#	178#	-0.01
34	Benzo(k)fluoranthene	1.392	1.350	3.0	121	0.00
35 C	Benzo(a)pyrene	0.850	1.112	-30.8#	178#	0.00
36	Dibenzo(a,h)anthracene	0.766	1.076	-40.5#	208#	-0.01
37	Benzo(g,h,i)perylene	0.923	1.192	-29.1#	171#	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1