

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081315\
 Data File : BE090512.D
 Acq On : 13 Aug 2015 22:28
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 14 04:35:45 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	138	0.00
2	1,4-Dioxane	0.687	0.550	19.9	103	0.00
3	n-Nitrosodimethylamine	0.773	0.727	6.0	130	0.00
4 S	2-Fluorophenol	0.750	1.334	-77.9#	263#	0.00
5 S	Phenol-d6	1.241	1.935	-55.9#	232#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	155#	0.00
7 S	Nitrobenzene-d5	0.331	0.330	0.3	163#	0.00
8	Nitrobenzene	0.349	0.343	1.7	170#	0.00
9	Naphthalene	1.104	0.999	9.5	141	0.00
10	Hexachlorobutadiene	0.191	0.144	24.6	116	0.00
11	2-Methylnaphthalene	0.640	0.631	1.4	160#	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	155#	0.00
13 S	2,4,6-Tribromophenol	0.095	0.170	-78.9#	355#	0.00
14 S	2-Fluorobiphenyl	1.375	1.263	8.1	140	0.00
15	Acenaphthylene	8.098	8.007	1.1	155#	0.00
16	Acenaphthene	1.259	1.218	3.3	150	0.00
17	Fluorene	1.635	1.484	9.2	139	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	144	0.00
19	4-Bromophenyl-phenylether	0.194	0.169	12.9	129	0.00
20	Hexachlorobenzene	1.005	0.769	23.5	110	0.00
21	Pentachlorophenol	0.044	0.062	-40.9#	328#	0.00
22	Phenanthrene	1.251	1.111	11.2	129	0.00
23	Anthracene	1.028	1.042	-1.4	150#	-0.02
24	Fluoranthene	1.231	1.197	2.8	146	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
26	Pyrene	1.084	1.247	-15.0	147	0.00
27 S	Terphenyl-d14	0.737	0.773	-4.9	130	-0.01
28	Benzo(a)anthracene	1.108	1.124	-1.4	139	0.00
29	Chrysene	1.339	1.124	16.1	101	0.00
30	Bis(2-ethylhexyl)phthalate	0.311	0.629	-102.3#	346#	0.00
31	Indeno(1,2,3-cd)pyrene	0.942	1.075	-14.1	170#	0.00
32 I	Perylene-d12	1.000	1.000	0.0	132	0.00
33	Benzo(b)fluoranthene	0.918	1.033	-12.5	158#	0.00
34	Benzo(k)fluoranthene	1.392	1.166	16.2	107	0.00
35 C	Benzo(a)pyrene	0.850	0.947	-11.4	157#	0.00
36	Dibenzo(a,h)anthracene	0.766	0.894	-16.7	178#	-0.01
37	Benzo(g,h,i)perylene	0.923	1.014	-9.9	150#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0