

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081315\
 Data File : BE090512.D
 Acq On : 13 Aug 2015 22:28
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 14 04:35:45 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE080515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 06 12:28:08 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	138	0.00
2	1,4-Dioxane	1.000	0.815	18.5	103	0.00
3	n-Nitrosodimethylamine	1.000	0.939	6.1	130	0.00
4 S	2-Fluorophenol	1.000	1.779	-77.9#	263	0.00
5 S	Phenol-d6	1.000	1.401	-40.1#	232	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	155	0.00
7 S	Nitrobenzene-d5	1.000	0.980	2.0	163	0.00
8	Nitrobenzene	1.000	0.964	3.6	170	0.00
9	Naphthalene	1.000	0.904	9.6	141	0.00
10	Hexachlorobutadiene	1.000	0.799	20.1	116	0.00
11	2-Methylnaphthalene	1.000	0.986	1.4	160	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	155	0.00
13 S	2,4,6-Tribromophenol	1.000	1.819	-81.9#	355	0.00
14 S	2-Fluorobiphenyl	1.000	0.919	8.1	140	0.00
15	Acenaphthylene	1.000	0.989	1.1	155	0.00
16	Acenaphthene	1.000	0.967	3.3	150	0.00
17	Fluorene	1.000	0.908	9.2	139	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	144	0.00
19	4-Bromophenyl-phenylether	1.000	0.871	12.9	129	0.00
20	Hexachlorobenzene	1.000	0.765	23.5	110	0.00
21	Pentachlorophenol	1.000	1.646	-64.6#	328	0.00
22	Phenanthrene	1.000	0.888	11.2	129	0.00
23	Anthracene	1.000	1.013	-1.3	150	-0.02
24	Fluoranthene	1.000	0.972	2.8	146	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	123	0.00
26	Pyrene	1.000	1.150	-15.0	147	0.00
27 S	Terphenyl-d14	1.000	1.049	-4.9	130	-0.01
28	Benzo(a)anthracene	1.000	1.014	-1.4	139	0.00
29	Chrysene	1.000	0.839	16.1	101	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	2.079	-107.9#	346	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.097	-9.7	170	0.00
32 I	Perylene-d12	5.000	5.000	0.0	132	0.00
33	Benzo(b)fluoranthene	1.000	1.071	-7.1	158	0.00
34	Benzo(k)fluoranthene	1.000	0.837	16.3	107	0.00
35 C	Benzo(a)pyrene	1.000	1.072	-7.2	157	0.00
36	Dibenzo(a,h)anthracene	1.000	1.089	-8.9	178	-0.01
37	Benzo(g,h,i)perylene	1.000	1.099	-9.9	150	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0