

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081415\
 Data File : BE090525.D
 Acq On : 14 Aug 2015 17:58
 Operator : UM/IZ
 Sample : G3248-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 CSB-02

Quant Time: Aug 17 15:29:39 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE081415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 14 16:52:42 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	21466	5.00	ng	0.00
6) Naphthalene-d8	10.33	136	104984	5.00	ng	0.00
12) Acenaphthene-d10	14.19	164	55594	5.00	ng	0.00
18) Phenanthrene-d10	16.95	188	123203	5.00	ng	0.00
25) Chrysene-d12	21.09	240	127026	5.00	ng	0.00
32) Perylene-d12	23.40	264	117453	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.22	112	13950	2.25	ng	0.00
5) Phenol-d6	6.77	99	28047	3.07	ng	0.00
7) Nitrobenzene-d5	8.72	82	14974	1.79	ng	0.00
13) 2,4,6-Tribromophenol	15.69	330	11241	5.25	ng	-0.02
14) 2-Fluorobiphenyl	12.82	172	39081	2.35	ng	0.00
27) Terphenyl-d14	19.55	244	71196	3.06	ng	0.00
Target Compounds						Qvalue
22) Phenanthrene	16.98	178	3458	0.11	ng	97
24) Fluoranthene	18.98	202	4100	0.12	ng	# 92
26) Pyrene	19.34	202	2016	0.06	ng	96
28) Benzo(a)anthracene	21.08	228	1162	0.03	ng	# 76
29) Chrysene	21.13	228	2327	0.07	ng	# 78
30) Bis(2-ethylhexyl)phthalate	21.02	149	7444	0.50	ng	99
31) Indeno(1,2,3-cd)pyrene	25.77	276	1712	0.05	ng	97
37) Benzo(g,h,i)perylene	26.51	276	1589	0.06	ng	# 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081415\
Data File : BE090525.D
Acq On : 14 Aug 2015 17:58
Operator : UM/IZ
Sample : G3248-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
CSB-02

Quant Time: Aug 17 15:29:39 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE081415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Aug 14 16:52:42 2015
Response via : Initial Calibration

