

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081415\
 Data File : BE090514.D
 Acq On : 14 Aug 2015 10:06
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 14 15:26:29 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE081415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 14 15:17:14 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	116	0.00
2	1,4-Dioxane	1.000	0.951	4.9	115	0.00
3	n-Nitrosodimethylamine	1.000	1.013	-1.3	124	0.00
4 S	2-Fluorophenol	1.000	0.959	4.1	118	0.00
5 S	Phenol-d6	1.000	0.994	0.6	123	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	117	0.00
7 S	Nitrobenzene-d5	1.000	1.011	-1.1	123	0.00
8	Nitrobenzene	1.000	1.021	-2.1	125	0.00
9	Naphthalene	1.000	0.938	6.2	117	0.00
10	Hexachlorobutadiene	1.000	0.934	6.6	116	0.00
11	2-Methylnaphthalene	1.000	0.981	1.9	123	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	121	0.00
13 S	2,4,6-Tribromophenol	1.000	1.025	-2.5	133	-0.02
14 S	2-Fluorobiphenyl	1.000	0.960	4.0	122	0.00
15	Acenaphthylene	1.000	0.967	3.3	124	0.00
16	Acenaphthene	1.000	0.943	5.7	122	0.00
17	Fluorene	1.000	0.957	4.3	123	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	122	0.00
19	4-Bromophenyl-phenylether	1.000	0.963	3.7	127	0.00
20	Hexachlorobenzene	1.000	0.922	7.8	120	0.00
21	Pentachlorophenol	1.000	1.109	-10.9	152	0.00
22	Phenanthrene	1.000	0.915	8.5	121	0.00
23	Anthracene	1.000	0.972	2.8	127	0.00
24	Fluoranthene	1.000	0.978	2.2	128	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	126	0.00
26	Pyrene	1.000	0.940	6.0	128	0.00
27 S	Terphenyl-d14	1.000	0.954	4.6	128	0.00
28	Benzo(a)anthracene	1.000	0.924	7.6	129	0.00
29	Chrysene	1.000	0.947	5.3	128	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.058	-5.8	159	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.973	2.7	135	0.00
32 I	Perylene-d12	5.000	5.000	0.0	129	0.00
33	Benzo(b)fluoranthene	1.000	0.978	2.2	135	0.00
34	Benzo(k)fluoranthene	1.000	0.963	3.7	129	0.00
35 C	Benzo(a)pyrene	1.000	0.986	1.4	134	0.00
36	Dibenzo(a,h)anthracene	1.000	0.983	1.7	135	0.00
37	Benzo(g,h,i)perylene	1.000	0.977	2.3	131	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0