

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081417\
 Data File : BE093786.D
 Acq On : 14 Aug 2017 14:17
 Operator : SJ/JU
 Sample : I4504-22
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AD2

Quant Time: Aug 14 18:14:02 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 12 02:01:53 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	2775	0.40	ng/ul	0.00
2) Naphthalene-d8	10.71	136	14666	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.52	164	8967	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	20900	0.40	ng/ul	0.00
16) Chrysene-d12	21.41	240	21368	0.40	ng/ul	0.00
20) Perylene-d12	23.91	264	20682	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.29	152	5398	0.23	ng/ul	0.00
14) Fluoranthene-d10	19.27	212	14246	0.22	ng/ul	0.00
Target Compounds				Qvalue		
5) 2-Methylnaphthalene	12.36	142	1141	0.044	ng/ul	99
9) Fluorene	15.57	166	1555	0.042	ng/ul#	88
12) Phenanthrene	17.29	178	3662	0.064	ng/ul#	94
13) Anthracene	17.29	178	3568	0.064	ng/ul	96
15) Fluoranthene	19.30	202	2068	0.029	ng/ul	85
17) Pyrene	19.66	202	2547	0.041	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081417\
Data File : BE093786.D
Acq On : 14 Aug 2017 14:17
Operator : SJ/JU
Sample : I4504-22
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AD2

Quant Time: Aug 14 18:14:02 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 12 02:01:53 2017
Response via : Initial Calibration

