

Data Path : Z:\HPCHEM1\BNA_E\Data\BE081417\
 Data File : BE093794.D
 Acq On : 14 Aug 2017 19:27
 Operator : SJ/JU
 Sample : I4557-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AC5

Quant Time: Aug 15 10:48:34 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 12 02:01:53 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	2399	0.40	ng/ul	0.00
2) Naphthalene-d8	10.70	136	12423	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.53	164	8613	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	16536	0.40	ng/ul	0.00
16) Chrysene-d12	21.41	240	17951	0.40	ng/ul	0.00
20) Perylene-d12	23.75	264	769138	0.40	ng/ul	-0.15
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.29	152	5956	0.30	ng/ul	0.00
14) Fluoranthene-d10	19.27	212	24550	0.47	ng/ul	0.00
Target Compounds					Qvalue	
3) Naphthalene	10.75	128	25888	0.857	ng/ul#	93
5) 2-Methylnaphthalene	12.36	142	43191	1.959	ng/ul	99
7) Acenaphthylene	14.25	152	3021	0.084	ng/ul#	48
8) Acenaphthene	14.59	153	6562	0.232	ng/ul	94
9) Fluorene	15.57	166	33139	0.925	ng/ul	89
12) Phenanthrene	17.29	178	58284	1.294	ng/ul#	90
13) Anthracene	17.29	178	57054	1.301	ng/ul	92
15) Fluoranthene	19.30	202	33444	0.594	ng/ul	83
17) Pyrene	19.66	202	45763	0.873	ng/ul#	85
18) Benzo(a)anthracene	21.39	228	15201	0.292	ng/ul#	74
19) Chrysene	21.45	228	14248	0.285	ng/ul#	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\Data\BE081417\
Data File : BE093794.D
Acq On : 14 Aug 2017 19:27
Operator : SJ/JU
Sample : I4557-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AC5

Quant Time: Aug 15 10:48:34 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 12 02:01:53 2017
Response via : Initial Calibration

