

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081517\
 Data File : BE093804.D
 Acq On : 15 Aug 2017 12:58
 Operator : SJ/JU
 Sample : PB101469BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SBLK69

Quant Time: Aug 16 05:45:49 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 16 05:35:21 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	2231	0.40	ng/ul	0.00
2) Naphthalene-d8	10.70	136	11924	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.53	164	7373	0.40	ng/ul	0.00
10) Phenanthrene-d10	0.00	188	0	0.00	ng/ul	-17.26
16) Chrysene-d12	21.41	240	15983	0.40	ng/ul	0.00
20) Perylene-d12	23.76	264	930945	0.40	ng/ul	-0.16
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.28	152	5587	0.29	ng/ul	0.00
14) Fluoranthene-d10	19.27	212	13142	0.00	ng/ul	0.00
Target Compounds						
9) Fluorene	15.51	166	1680	0.055	ng/ul#	22
17) Pyrene	19.66	202	1105	0.024	ng/ul#	82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081517\
Data File : BE093804.D
Acq On : 15 Aug 2017 12:58
Operator : SJ/JU
Sample : PB101469BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SBLK69

Quant Time: Aug 16 05:45:49 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE072417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 16 05:35:21 2017
Response via : Initial Calibration





#9

Fluorene

Concen: 0.055 ng/ul

RT: 15.51 min Scan# 2293

Delta R.T. -0.06 min

Lab File: BE093804.D

Acq: 15 Aug 2017 12:58

Instrument :

BNA_E

ClientSampleId :

SBLK69





Tot Ion:166 Resp: 1680

Ion Ratio Lower Upper

166 100

165 7.9 73.4 110.2#

167 31.8 14.6 22.0#



#17

Pyrene

Concen: 0.024 ng/ul

RT: 19.66 min Scan# 2831

Delta R.T. -0.00 min

Lab File: BE093804.D

Acq: 15 Aug 2017 12:58

Tgt Ion:202 Resp: 1105

Ion Ratio Lower Upper

202 100

101 22.7 18.2 27.4

100 37.1 15.6 23.4#

