

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081517\
 Data File : BE093820.D
 Acq On : 15 Aug 2017 22:52
 Operator : SJ/JU
 Sample : I4672-14
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AJ2

Manual Integrations
 APPROVED

Sohil
 8/16/2017 6:50:01 PM

Quant Time: Aug 16 15:28:35 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 16 07:30:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	2604	0.40	ng/ul	0.00
2) Naphthalene-d8	10.70	136	14050	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.52	164	8534	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	17811	0.40	ng/ul	0.00
16) Chrysene-d12	21.41	240	18454	0.40	ng/ul	0.00
20) Perylene-d12	23.91	264	13983m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.28	152	5425	0.24	ng/ul	0.00
14) Fluoranthene-d10	19.27	212	13377	0.24	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.76	128	5427	0.159	ng/ul#	94
5) 2-Methylnaphthalene	12.36	142	2245	0.090	ng/ul	99
7) Acenaphthylene	14.24	152	1307	0.037	ng/ul#	84
8) Acenaphthene	14.59	153	2129	0.076	ng/ul	99
9) Fluorene	15.57	166	2518	0.071	ng/ul	94
12) Phenanthrene	17.29	178	35066	0.723	ng/ul#	89
13) Anthracene	17.38	178	9489	0.201	ng/ul#	91
15) Fluoranthene	19.30	202	76674	1.264	ng/ul	84
17) Pyrene	19.66	202	71846	1.334	ng/ul#	83
18) Benzo(a)anthracene	21.39	228	38115	0.712	ng/ul#	87
19) Chrysene	21.45	228	36646	0.714	ng/ul	99
21) Benzo(b)fluoranthene	23.15	252	36274	0.839	ng/ul	89
22) Benzo(k)fluoranthene	23.17	252	17079m	0.400	ng/ul	
23) Benzo(a)pyrene	23.80	252	30714m	0.747	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.55	276	14726m	0.324	ng/ul	
25) Dibenzo(a,h)anthracene	26.56	278	3771m	0.099	ng/ul	
26) Benzo(g,h,i)perylene	27.37	276	11650m	0.304	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081517\
Data File : BE093820.D
Acq On : 15 Aug 2017 22:52
Operator : SJ/JU
Sample : I4672-14
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AJ2

Manual Integrations
APPROVED

Sohil
8/16/2017 6:50:01 PM

Quant Time: Aug 16 15:28:35 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 16 07:30:49 2017
Response via : Initial Calibration

