

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081517\
 Data File : BE093824.D
 Acq On : 16 Aug 2017 1:21
 Operator : SJ/JU
 Sample : I4672-23
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AH6

Quant Time: Aug 16 08:19:53 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 16 07:30:49 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	2356	0.40	ng/ul	0.00
2) Naphthalene-d8	10.70	136	12749	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.52	164	7760	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	16836	0.40	ng/ul	0.00
16) Chrysene-d12	21.41	240	18379	0.40	ng/ul	0.00
20) Perylene-d12	23.91	264	22385	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.29	152	2721	0.13	ng/ul	0.00
14) Fluoranthene-d10	19.27	212	7266	0.14	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.76	128	4252	0.137	ng/ul#	96
5) 2-Methylnaphthalene	12.36	142	4928	0.218	ng/ul	99
12) Phenanthrene	17.29	178	11927	0.260	ng/ul#	92
15) Fluoranthene	19.30	202	2234	0.039	ng/ul	85
17) Pyrene	19.66	202	3918	0.073	ng/ul	95
19) Chrysene	21.45	228	2815	0.055	ng/ul#	84
26) Benzo(g,h,i)perylene	27.37	276	1874	0.031	ng/ul#	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081517\
Data File : BE093824.D
Acq On : 16 Aug 2017 1:21
Operator : SJ/JU
Sample : I4672-23
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AH6

Quant Time: Aug 16 08:19:53 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 16 07:30:49 2017
Response via : Initial Calibration

