

Data Path : Z:\HPCHEM1\BNA_E\Data\BE081517\
 Data File : BE093802.D
 Acq On : 15 Aug 2017 11:29
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.406

Quant Time: Aug 16 05:31:04 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 15 10:51:12 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	70	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	72	0.00
3	Naphthalene	0.973	0.961	1.2	73	0.00
4 SURR	2-Methylnaphthalene-d10	0.648	0.619	4.5	70	0.00
5	2-Methylnaphthalene	0.710	0.656	7.6	68	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	62	0.00
7	Acenaphthylene	1.673	1.720	-2.8	65	0.00
8 C	Acenaphthene	1.316	1.286	2.3	61	0.00
9	Fluorene	1.664	1.480	11.1	56	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	52	0.00
11	Pentachlorophenol	0.066	0.068	-3.0	61	0.00
12	Phenanthrene	1.089	1.055	3.1	53	0.00
13	Anthracene	1.061	1.033	2.6	53	0.00
14 SURR	Fluoranthene-d10	1.263	1.168	7.5	50	0.00
15 C	Fluoranthene	1.362	1.236	9.3	50	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	52	0.00
17	Pyrene	1.168	1.133	3.0	51	0.00
18	Benzo(a)anthracene	1.160	1.112	4.1	51	0.00
19	Chrysene	1.112	1.101	1.0	52	0.00
20 I	Perylene-d12	1.000	1.000	0.0	49	0.00
21	Benzo(b)fluoranthene	1.237	1.149	7.1	48	0.00
22	Benzo(k)fluoranthene	1.220	1.231	-0.9	52	0.00
23 C	Benzo(a)pyrene	1.177	1.159	1.5	51	0.00
24	Indeno(1,2,3-cd)pyrene	1.299	1.145	11.9	46	0.00
25	Dibenzo(a,h)anthracene	1.086	0.986	9.2	47	0.00
26	Benzo(g,h,i)perylene	1.095	0.901	17.7	42	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0