

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081517\
 Data File : BE093831.D
 Acq On : 16 Aug 2017 5:44
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.408

Quant Time: Aug 16 07:39:24 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 16 07:30:49 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	65	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	67	0.00
3	Naphthalene	0.973	0.961	1.2	68	0.00
4 SURR	2-Methylnaphthalene-d10	0.648	0.625	3.5	66	0.00
5	2-Methylnaphthalene	0.710	0.664	6.5	64	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	59	0.00
7	Acenaphthylene	1.673	1.745	-4.3	64	0.00
8 C	Acenaphthene	1.316	1.281	2.7	59	0.00
9	Fluorene	1.664	1.479	11.1	54	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	50	0.00
11	Pentachlorophenol	0.066	0.070	-6.1	60	0.00
12	Phenanthrene	1.089	1.048	3.8	50	0.00
13	Anthracene	1.061	1.050	1.0	52	0.00
14 SURR	Fluoranthene-d10	1.263	1.185	6.2	49	0.00
15 C	Fluoranthene	1.362	1.238	9.1	48	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	48	0.00
17	Pyrene	1.168	1.182	-1.2	49	0.00
18	Benzo(a)anthracene	1.160	1.123	3.2	48	0.00
19	Chrysene	1.112	1.113	-0.1	49	0.00
20 I	Perylene-d12	1.000	1.000	0.0	46	0.00
21	Benzo(b)fluoranthene	1.237	1.138	8.0	44	0.00
22	Benzo(k)fluoranthene	1.220	1.234	-1.1	48	0.00
23 C	Benzo(a)pyrene	1.177	1.161	1.4	47	0.00
24	Indeno(1,2,3-cd)pyrene	1.299	1.124	13.5	41	0.00
25	Dibenzo(a,h)anthracene	1.086	0.937	13.7	42	0.00
26	Benzo(g,h,i)perylene	1.095	0.918	16.2	40	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0