

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081517\
 Data File : BE093815.D
 Acq On : 15 Aug 2017 19:47
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.407

Quant Time: Aug 16 05:46:52 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 16 05:35:21 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	2107	0.40	ng/ul	0.00
2) Naphthalene-d8	10.71	136	11061	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.53	164	6084	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	13847	0.40	ng/ul	0.00
16) Chrysene-d12	21.41	240	14836	0.40	ng/ul	0.00
20) Perylene-d12	23.91	264	12375	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.29	152	6869	0.38	ng/ul	0.00
14) Fluoranthene-d10	19.28	212	15545	0.36	ng/ul	0.00
Target Compounds					Qvalue	
3) Naphthalene	10.75	128	10569	0.393	ng/ul#	92
5) 2-Methylnaphthalene	12.36	142	7220	0.368	ng/ul	99
7) Acenaphthylene	14.25	152	10297	0.405	ng/ul	96
8) Acenaphthene	14.59	153	7836	0.391	ng/ul	96
9) Fluorene	15.57	166	8892	0.351	ng/ul#	95
11) Pentachlorophenol	16.93	266	878	0.384	ng/ul	94
12) Phenanthrene	17.29	178	14474	0.384	ng/ul#	89
13) Anthracene	17.38	178	14112	0.384	ng/ul#	87
15) Fluoranthene	19.30	202	16298	0.346	ng/ul	85
17) Pyrene	19.66	202	17144	0.396	ng/ul#	84
18) Benzo(a)anthracene	21.40	228	16059	0.373	ng/ul#	86
19) Chrysene	21.45	228	16520	0.400	ng/ul	99
21) Benzo(b)fluoranthene	23.14	252	14634	0.382	ng/ul	87
22) Benzo(k)fluoranthene	23.20	252	15173	0.402	ng/ul	90
23) Benzo(a)pyrene	23.80	252	14550	0.400	ng/ul#	85
24) Indeno(1,2,3-cd)pyrene	26.57	276	13920	0.346	ng/ul#	85
25) Dibenzo(a,h)anthracene	26.59	278	11541	0.344	ng/ul	94
26) Benzo(g,h,i)perylene	27.38	276	11158	0.329	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081517\
Data File : BE093815.D
Acq On : 15 Aug 2017 19:47
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.407

Quant Time: Aug 16 05:46:52 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE072417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 16 05:35:21 2017
Response via : Initial Calibration

