

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081617\
 Data File : BE093858.D
 Acq On : 17 Aug 2017 2:16
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.411

Quant Time: Aug 17 03:08:21 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE072417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 17 02:52:40 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	69	0.00
3	Naphthalene	0.973	0.965	0.8	70	0.00
4 SURR	2-Methylnaphthalene-d10	0.648	0.611	5.7	66	0.00
5	2-Methylnaphthalene	0.710	0.648	8.7	64	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	59	0.00
7	Acenaphthylene	1.673	1.767	-5.6	64	0.00
8 C	Acenaphthene	1.316	1.273	3.3	58	0.00
9	Fluorene	1.664	1.452	12.7	52	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	48	0.00
11	Pentachlorophenol	0.066	0.068	-3.0	55	0.00
12	Phenanthrene	1.089	1.037	4.8	47	0.00
13	Anthracene	1.061	1.052	0.8	49	0.00
14 SURR	Fluoranthene-d10	1.263	1.174	7.0	46	0.00
15 C	Fluoranthene	1.362	1.235	9.3	45	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	44	0.00
17	Pyrene	1.168	1.225	-4.9	47	0.00
18	Benzo(a)anthracene	1.160	1.142	1.6	45	0.00
19	Chrysene	1.112	1.117	-0.4	45	0.00
20 I	Perylene-d12	1.000	1.000	0.0	45	0.00
21	Benzo(b)fluoranthene	1.237	1.127	8.9	43	0.00
22	Benzo(k)fluoranthene	1.220	1.176	3.6	45	0.00
23 C	Benzo(a)pyrene	1.177	1.134	3.7	45	0.00
24	Indeno(1,2,3-cd)pyrene	1.299	1.234	5.0	45	0.00
25	Dibenzo(a,h)anthracene	1.086	1.009	7.1	44	0.00
26	Benzo(a,h,i)perylene	1.095	1.069	2.4	46	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0