

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081715\
 Data File : BE090532.D
 Acq On : 17 Aug 2015 13:42
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 18 04:34:14 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE081415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 14 16:52:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.658	0.651	1.1	100	0.00
3	n-Nitrosodimethylamine	0.863	0.853	1.2	108	0.00
4 S	2-Fluorophenol	1.446	1.494	-3.3	110	0.00
5 S	Phenol-d6	2.125	2.210	-4.0	116	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
7 S	Nitrobenzene-d5	0.398	0.417	-4.8	119	0.00
8	Nitrobenzene	0.413	0.429	-3.9	120	0.00
9	Naphthalene	1.136	1.120	1.4	107	0.00
10	Hexachlorobutadiene	0.167	0.163	2.4	106	0.00
11	2-Methylnaphthalene	0.740	0.766	-3.5	115	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
13 S	2,4,6-Tribromophenol	0.192	0.212	-10.4	139	0.00
14 S	2-Fluorobiphenyl	1.493	1.486	0.5	114	0.00
15	Acenaphthylene	9.132	9.215	-0.9	118	0.00
16	Acenaphthene	1.374	1.366	0.6	115	0.00
17	Fluorene	1.684	1.713	-1.7	117	0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
19	4-Bromophenyl-phenylether	0.204	0.209	-2.5	122	0.00
20	Hexachlorobenzene	0.906	0.902	0.4	116	0.00
21	Pentachlorophenol	0.079	0.063	20.3	123	0.00
22	Phenanthrene	1.312	1.295	1.3	117	0.00
23	Anthracene	1.208	1.226	-1.5	122	0.00
24	Fluoranthene	1.396	1.422	-1.9	121	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
26	Pyrene	1.430	1.436	-0.4	124	0.00
27 S	Terphenyl-d14	0.914	0.946	-3.5	126	0.00
28	Benzo(a)anthracene	1.316	1.332	-1.2	128	0.00
29	Chrysene	1.344	1.362	-1.3	123	0.00
30	Bis(2-ethylhexyl)phthalate	0.725	0.963	-32.8#	194#	0.00
31	Indeno(1,2,3-cd)pyrene	1.266	1.323	-4.5	136	0.00
32 I	Perylene-d12	1.000	1.000	0.0	126	0.00
33	Benzo(b)fluoranthene	1.184	1.222	-3.2	133	0.00
34	Benzo(k)fluoranthene	1.342	1.403	-4.5	131	0.00
35 C	Benzo(a)pyrene	1.108	1.175	-6.0	138	0.00
36	Dibenzo(a,h)anthracene	1.069	1.126	-5.3	140	0.00
37	Benzo(g,h,i)perylene	1.169	1.231	-5.3	135	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0