

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE081715\
 Data File : BE090548.D
 Acq On : 18 Aug 2015 00:24
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Aug 18 06:12:00 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE081415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Aug 14 16:52:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.658	0.638	3.0	103	0.00
3	n-Nitrosodimethylamine	0.863	0.818	5.2	108	0.00
4 S	2-Fluorophenol	1.446	1.544	-6.8	119	0.00
5 S	Phenol-d6	2.125	2.241	-5.5	123	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
7 S	Nitrobenzene-d5	0.398	0.398	0.0	121	0.00
8	Nitrobenzene	0.413	0.409	1.0	122	0.00
9	Naphthalene	1.136	1.115	1.8	114	0.00
10	Hexachlorobutadiene	0.167	0.163	2.4	112	0.00
11	2-Methylnaphthalene	0.740	0.754	-1.9	120	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
13 S	2,4,6-Tribromophenol	0.192	0.198	-3.1	134	0.00
14 S	2-Fluorobiphenyl	1.493	1.496	-0.2	119	0.00
15	Acenaphthylene	9.132	9.156	-0.3	121	0.00
16	Acenaphthene	1.374	1.370	0.3	120	0.00
17	Fluorene	1.684	1.662	1.3	118	0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
19	4-Bromophenyl-phenylether	0.204	0.209	-2.5	122	0.00
20	Hexachlorobenzene	0.906	0.912	-0.7	117	0.00
21	Pentachlorophenol	0.079	0.076	3.8	148	0.00
22	Phenanthrene	1.312	1.278	2.6	115	0.00
23	Anthracene	1.208	1.213	-0.4	120	0.00
24	Fluoranthene	1.396	1.372	1.7	117	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
26	Pyrene	1.430	1.443	-0.9	118	0.00
27 S	Terphenyl-d14	0.914	0.936	-2.4	118	0.00
28	Benzo(a)anthracene	1.316	1.288	2.1	117	0.00
29	Chrysene	1.344	1.344	0.0	115	0.00
30	Bis(2-ethylhexyl)phthalate	0.725	0.814	-12.3	155#	0.00
31	Indeno(1,2,3-cd)pyrene	1.266	1.265	0.1	123	0.00
32 I	Perylene-d12	1.000	1.000	0.0	118	0.00
33	Benzo(b)fluoranthene	1.184	1.201	-1.4	123	0.00
34	Benzo(k)fluoranthene	1.342	1.367	-1.9	119	0.00
35 C	Benzo(a)pyrene	1.108	1.124	-1.4	124	0.00
36	Dibenzo(a,h)anthracene	1.069	1.071	-0.2	125	0.00
37	Benzo(g,h,i)perylene	1.169	1.200	-2.7	123	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0