


Data Path : Z:\HPCHEM1\BNA\_E\Data\BE081815\  
 Data File : BE090552.D  
 Acq On : 18 Aug 2015 13:02  
 Operator : UM/IZ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_E\METHODS\8270-SIM-BE081415.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri Aug 14 16:52:42 2015



AutoFind: Scans 2517, 2518, 2519; Background Corrected with Scan 2508

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 51.9      | 30038   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.2       | 76      | PASS             |
| 69          | 198          | 0.00         | 100          | 52.6      | 30426   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.4       | 116     | PASS             |
| 127         | 198          | 10           | 80           | 57.1      | 33025   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 57871   | PASS             |
| 199         | 198          | 5            | 9            | 7.2       | 4188    | PASS             |
| 275         | 198          | 10           | 60           | 25.7      | 14883   | PASS             |
| 365         | 198          | 1            | 100          | 3.3       | 1906    | PASS             |
| 441         | 442          | 0.01         | 24           | 16.7      | 9602    | PASS             |
| 442         | 198          | 50           | 100          | 99.1      | 57334   | PASS             |
| 443         | 442          | 15           | 24           | 19.6      | 11209   | PASS             |