

Data Path : Z:\HPCHEM1\BNA E\Data\BE081916\
 Data File : BE092269.D
 Acq On : 19 Aug 2016 14:11
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 22 00:56:21 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 10 11:12:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	122	0.02
2	1,4-Dioxane	0.638	0.686	-7.5	122	0.01
3	n-Nitrosodimethylamine	0.775	0.776	-0.1	137	0.00
4 S	2-Fluorophenol	1.094	1.185	-8.3	145	0.00
5 S	Phenol-d6	1.573	1.519	3.4	150#	0.02
6	bis(2-Chloroethyl)ether	1.275	1.234	3.2	131	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
8 S	Nitrobenzene-d5	0.353	0.344	2.5	141	0.00
9	Naphthalene	1.121	1.143	-2.0	125	0.00
10	Hexachlorobutadiene	0.162	0.167	-3.1	126	0.00
11	2-Methylnaphthalene	0.700	0.728	-4.0	129	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
13 S	2,4,6-Tribromophenol	0.150	0.131	12.7	137	0.00
14 S	2-Fluorobiphenyl	1.545	1.580	-2.3	127	-0.01
15	Acenaphthylene	8.647	8.629	0.2	125	0.00
16	Acenaphthene	1.388	1.420	-2.3	125	0.00
17	Fluorene	1.677	1.735	-3.5	128	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.02
19	Hexachlorobenzene	0.198	0.194	2.0	120	0.02
20	Pentachlorophenol	0.042	0.067	-59.5#	242#	0.02
21	Phenanthrene	1.293	1.232	4.7	122	0.02
22	Anthracene	1.169	1.133	3.1	129	0.02
23	Fluoranthene	5.137	5.028	2.1	127	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
25	Pyrene	5.914	6.174	-4.4	129	0.00
26 S	Terphenyl-d14	0.845	0.892	-5.6	135	0.00
27	Benzo(a)anthracene	6.002	6.383	-6.3	135	0.02
28	Chrysene	5.546	5.597	-0.9	117	0.02
29	Bis(2-ethylhexyl)phthalate	0.770	0.746	3.1	140	0.02
30	Indeno(1,2,3-cd)pyrene	5.972	5.596	6.3	122	0.00
31 I	Perylene-d12	1.000	1.000	0.0	122	0.00
32	Benzo(b)fluoranthene	1.313	1.259	4.1	127	0.00
33	Benzo(k)fluoranthene	1.323	1.373	-3.8	125	0.00
34 C	Benzo(a)pyrene	1.238	1.206	2.6	129	0.00
35	Dibenzo(a,h)anthracene	1.151	0.996	13.5	118	0.00
36	Benzo(g,h,i)perylene	4.954	4.748	4.2	123	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0