

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081916\
 Data File : BE092274.D
 Acq On : 19 Aug 2016 17:37
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 22 01:03:22 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 10 11:12:32 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.02
2	1,4-Dioxane	0.638	0.721	-13.0	118	0.00
3	n-Nitrosodimethylamine	0.775	0.775	0.0	125	0.00
4 S	2-Fluorophenol	1.094	1.056	3.5	119	0.00
5 S	Phenol-d6	1.573	1.345	14.5	122	0.02
6	bis(2-Chloroethyl)ether	1.275	1.257	1.4	122	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
8 S	Nitrobenzene-d5	0.353	0.356	-0.8	134	0.00
9	Naphthalene	1.121	1.152	-2.8	116	-0.02
10	Hexachlorobutadiene	0.162	0.172	-6.2	120	0.00
11	2-Methylnaphthalene	0.700	0.726	-3.7	119	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
13 S	2,4,6-Tribromophenol	0.150	0.127	15.3	121	0.00
14 S	2-Fluorobiphenyl	1.545	1.596	-3.3	117	-0.01
15	Acenaphthylene	8.647	8.617	0.3	114	0.00
16	Acenaphthene	1.388	1.435	-3.4	116	0.00
17	Fluorene	1.677	1.727	-3.0	116	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.02
19	Hexachlorobenzene	0.198	0.201	-1.5	116	0.02
20	Pentachlorophenol	0.042	0.035	16.7	117	0.02
21	Phenanthrene	1.293	1.304	-0.9	119	0.02
22	Anthracene	1.169	1.104	5.6	117	0.02
23	Fluoranthene	5.137	4.866	5.3	114	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	114	0.02
25	Pyrene	5.914	5.911	0.1	119	0.00
26 S	Terphenyl-d14	0.845	0.825	2.4	120	0.00
27	Benzo(a)anthracene	6.002	5.832	2.8	119	0.02
28	Chrysene	5.546	5.596	-0.9	113	0.02
29	Bis(2-ethylhexyl)phthalate	0.770	0.572	25.7#	104	0.02
30	Indeno(1,2,3-cd)pyrene	5.972	5.460	8.6	115	0.00
31 I	Perylene-d12	1.000	1.000	0.0	111	0.00
32	Benzo(b)fluoranthene	1.313	1.251	4.7	115	0.00
33	Benzo(k)fluoranthene	1.323	1.364	-3.1	114	0.00
34 C	Benzo(a)pyrene	1.238	1.176	5.0	115	0.00
35	Dibenzo(a,h)anthracene	1.151	1.062	7.7	114	0.00
36	Benzo(g,h,i)perylene	4.954	4.785	3.4	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0