

Data Path : Z:\HPCHEM1\BNA E\DATA\BE081916\
 Data File : BE092274.D
 Acq On : 19 Aug 2016 17:37
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

sohil
 8/22/2016 3:50:09 PM

Quant Time: Aug 22 01:03:22 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 10 11:12:32 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.19	152	10366	5.00	ng	0.02
7) Naphthalene-d8	10.97	136	50053	5.00	ng	0.00
12) Acenaphthene-d10	14.83	164	28576	5.00	ng	0.00
18) Phenanthrene-d10	17.58	188	72093	5.00	ng	0.02
24) Chrysene-d12	21.77	240	64216	5.00	ng	0.02
31) Perylene-d12	24.13	264	65243	5.00	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.71	112	876	0.39	ng	0.00
5) Phenol-d6	7.38	99	1115	0.34	ng	0.02
8) Nitrobenzene-d5	9.36	82	1425m	0.40	ng	0.00
13) 2,4,6-Tribromophenol	16.33	330	290	0.50	ng	0.00
14) 2-Fluorobiphenyl	13.46	172	3648	0.41	ng	-0.01
26) Terphenyl-d14	20.19	244	4237	0.39	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.44	88	598	0.44	ng	95
3) n-Nitrosodimethylamine	3.81	42	643	0.48	ng	# 95
6) bis(2-Chloroethyl)ether	7.61	93	1042	0.39	ng	99
9) Naphthalene	11.01	128	4611	0.41	ng	# 94
10) Hexachlorobutadiene	11.32	225	687	0.42	ng	98
11) 2-Methylnaphthalene	12.66	142	2906	0.41	ng	95
15) Acenaphthylene	14.54	152	19700	0.40	ng	99
16) Acenaphthene	14.90	154	3281	0.41	ng	99
17) Fluorene	15.88	166	3949	0.41	ng	98
19) Hexachlorobenzene	16.91	284	1162m	0.41	ng	
20) Pentachlorophenol	17.26	266	202	0.38	ng	97
21) Phenanthrene	17.63	178	7518	0.40	ng	99
22) Anthracene	17.72	178	6367	0.38	ng	100
23) Fluoranthene	19.63	202	28062	0.38	ng	99
25) Pyrene	19.99	202	30367	0.40	ng	100
27) Benzo(a)anthracene	21.75	228	29963	0.39	ng	99
28) Chrysene	21.79	228	28749m	0.40	ng	
29) Bis(2-ethylhexyl)phthalate	21.66	149	2938	0.45	ng	# 98
30) Indeno(1,2,3-cd)pyrene	26.65	276	28051	0.37	ng	100
32) Benzo(b)fluoranthene	23.39	252	6530	0.38	ng	99
33) Benzo(k)fluoranthene	23.44	252	7118	0.41	ng	98
34) Benzo(a)pyrene	24.02	252	6140	0.38	ng	98
35) Dibenzo(a,h)anthracene	26.67	278	5545	0.37	ng	100
36) Benzo(g,h,i)perylene	27.42	276	24974	0.39	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE081916\
Data File : BE092274.D
Acq On : 19 Aug 2016 17:37
Operator : UM/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM

Quant Time: Aug 22 01:03:22 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 10 11:12:32 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 8.19 min Scan# 577

Delta R.T. 0.02 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC




Tot Ion:	152	Resp:	10366
Ion	Ratio	Lower	Upper
152	100		
150	167.8	135.6	203.4
115	70.2	59.4	89.0

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.44 ng

RT: 3.44 min Scan# 74




Delta R.T. 0.00 min


Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Tgt Ion:	88	Resp:	598
Ion	Ratio	Lower	Upper
88	100		
58	79.3	64.4	96.6
43	35.1	33.4	50.2

Abundance Ion 88.00 (87.70 to 88.70): BE0
Ion 58.00 (57.70 to 58.70): BE0
Ion 43.00 (42.70 to 43.70): BE0





#3





n-Nitrosodimethylamine
Concen: 0.48 ng
RT: 3.81 min Scan# 106
Delta R.T. 0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
42	100		
74	100.5	77.2	115.8
44	0.0	5.0	7.4#

Manual Integrations
APPROVED




sohil
8/22/2016 3:50:09 PM




#4

2-Fluorophenol
Concen: 0.39 ng
RT: 5.71 min Scan# 369
Delta R.T. -0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Tgt Ion	Ratio	Lower	Upper
112	100		
64	67.9	53.2	79.8
63	36.6	27.4	41.2





#5

Phenol-d6

Concen: 0.34 ng

RT: 7.38 min Scan# 542

Delta R.T. 0.02 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC




Tot Ion	Ratio	Resp	Lower	Upper
99	100	1115		
42	0.0	0.0	0.0	
71	0.0	29.8	44.8	

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM



Abundance Ion 99.00 (98.70 to 99.70): BE0
Ion 42.00 (41.70 to 42.70): BE0
Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 0.39 ng


RT: 7.61 min Scan# 552

Delta R.T. 0.02 min



Lab File: BE092274.D


Acq: 19 Aug 2016 17:37

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	1042		
63	85.7	68.5	102.7	
95	33.0	27.2	40.8	



Abundance Ion 93.00 (92.70 to 93.70): BE0
Ion 63.00 (62.70 to 63.70): BE0
Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.97 min Scan# 709

Delta R.T. -0.00 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tot Ion	Ratio	Resp	Lower	Upper
136	100			
137	11.1	8.8	13.2	
54	7.3	6.2	9.4	
68	5.3	4.6	6.8	

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM



#8

Nitrobenzene-d5

Concen: 0.40 ng m




RT: 9.36 min Scan# 628


Delta R.T. -0.00 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Tgt Ion	Ratio	Resp	Lower	Upper
82	100			
128	25.1	45.0	67.4#	
54	74.7	45.4	68.2#	





#9

Naphthalene

Concen: 0.41 ng

RT: 11.01 min Scan# 711

Delta R.T. -0.02 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC




Tot Ion:	128	Resp:	4611
Ion Ratio	Lower	Upper	
128	100		
129	12.4	11.5	17.3
127	16.6	11.0	16.6#

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E



#10

Hexachlorobutadiene

Concen: 0.42 ng

RT: 11.32 min Scan# 727




Delta R.T. -0.00 min


Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Tgt Ion:	225	Resp:	687
Ion Ratio	Lower	Upper	
225	100		
223	62.6	52.2	78.4
227	63.6	51.0	76.6

Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E






#11
2-Methylnaphthalene
Concen: 0.41 ng
RT: 12.66 min Scan# 817
Delta R.T. -0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
142	100		
141	83.9	70.6	105.8
115	35.3	31.0	46.4

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.83 min Scan# 981
Delta R.T. -0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37




Tgt Ion	Ratio	Lower	Upper
164	100		
162	103.0	83.3	124.9
160	45.9	38.0	57.0


Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#13
 2,4,6-Tribromophenol
 Concen: 0.50 ng
 RT: 16.33 min Scan# 1090
 Delta R.T. 0.00 min
 Lab File: BE092274.D
 Acq: 19 Aug 2016 17:37

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tot Ion:330 Resp: 290
 Ion Ratio Lower Upper
 330 100
 332 93.8 75.4 113.0
 141 44.1 37.7 56.5




Manual Integrations
 APPROVED


sohil
 8/22/2016 3:50:09 PM



#14
 2-Fluorobiphenyl
 Concen: 0.41 ng
 RT: 13.46 min Scan# 886
 Delta R.T. -0.01 min
 Lab File: BE092274.D
 Acq: 19 Aug 2016 17:37

Tgt Ion:172 Resp: 3648
 Ion Ratio Lower Upper
 172 100
 171 37.9 26.8 40.2
 170 27.1 17.8 26.6#





#15

Acenaphthylene

Concen: 0.40 ng

RT: 14.54 min Scan# 964

Delta R.T. -0.00 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tot Ion:	152	Resp:	19700
Ion Ratio	Lower	Upper	
152	100		
151	4.9	3.8	5.8
153	13.1	10.2	15.2

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM



#16

Acenaphthene

Concen: 0.41 ng




RT: 14.90 min Scan# 985


Delta R.T. -0.00 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Tgt Ion:	154	Resp:	3281
Ion Ratio	Lower	Upper	
154	100		
153	440.2	355.0	532.4
152	224.4	179.7	269.5





#17

Fluorene

Concen: 0.41 ng

RT: 15.88 min Scan# 1051

Delta R.T. 0.00 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tot Ion:	166	Resp:	3949
Ion	Ratio	Lower	Upper
166	100		
165	98.0	80.1	120.1
167	13.6	10.6	16.0

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.58 min Scan# 1154


Delta R.T. 0.02 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Tgt Ion:	188	Resp:	72093
Ion	Ratio	Lower	Upper
188	100		
94	10.0	9.2	13.8
80	9.9	9.1	13.7









#19
Hexachlorobenzene
Concen: 0.41 ng/ml
RT: 16.91 min Scan# 1125
Delta R.T. 0.02 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion:	284	Resp:	1162
Ion Ratio	Lower	Upper	
284	100		
142	0.0	0.0	0.0
249	0.0	0.0	0.0




Manual Integrations
APPROVED




sohil
8/22/2016 3:50:09 PM



#20
Pentachlorophenol
Concen: 0.38 ng/ml
RT: 17.26 min Scan# 1140
Delta R.T. 0.02 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Tgt Ion:	266	Resp:	202
Ion Ratio	Lower	Upper	
266	100		
268	74.8	62.3	93.5
264	82.2	67.4	101.0





#21

Phenanthrene

Concen: 0.40 ng

RT: 17.63 min Scan# 1156

Delta R.T. 0.02 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tot Ion:178	Resp:	7518
Ion Ratio	Lower	Upper
178 100		
176 19.2	15.1	22.7
179 16.3	12.9	19.3

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM



#22

Anthracene

Concen: 0.38 ng


RT: 17.72 min Scan# 1160


Delta R.T. 0.02 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Tgt Ion:178	Resp:	6367
Ion Ratio	Lower	Upper
178 100		
176 18.5	15.0	22.4
179 14.9	12.0	18.0





#23

Fluoranthene

Concen: 0.38 ng

RT: 19.63 min Scan# 1267

Delta R.T. -0.00 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Instrument :

BNA_E


ClientSampleId :

SSTDCCC0.4EC




Tot Ion:	202	Resp:	28062
Ion Ratio	Lower	Upper	
202	100		
101	2.8	2.2	3.4
203	17.4	13.7	20.5

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E



#24

Chrysene-d12

Concen: 5.00 ng


RT: 21.77 min Scan# 1376

Delta R.T. 0.02 min



Lab File: BE092274.D


Acq: 19 Aug 2016 17:37

Tgt Ion:	240	Resp:	64216
Ion Ratio	Lower	Upper	
240	100		
120	1.3	1.2	1.8
236	21.0	17.1	25.7



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E









#25
Pvrene
Concen: 0.40 ng
RT: 19.99 min Scan# 1288
Delta R.T. -0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
202	100		
200	5.4	4.2	6.4
203	17.5	14.0	21.0




Manual Integrations
APPROVED


sohil
8/22/2016 3:50:09 PM



#26
Terphenyl-d14
Concen: 0.39 ng
RT: 20.19 min Scan# 1300
Delta R.T. -0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.2	10.2	15.2
122	12.7	15.9	23.9#





#27

Benzo(a)anthracene

Concen: 0.39 ng

RT: 21.75 min Scan# 1375

Delta R.T. 0.02 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tot Ion:	228	Resp:	29963
Ion Ratio	Lower	Upper	
228	100		
226	19.1	15.9	23.9
229	27.3	22.4	33.6

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM



#28

Chrysene

Concen: 0.40 ng m




RT: 21.79 min Scan# 1377


Delta R.T. 0.02 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Tgt Ion:	228	Resp:	28749
Ion Ratio	Lower	Upper	
228	100		
226	23.0	18.3	27.5
229	23.3	18.6	27.8





#29

Bis(2-ethylhexyl)phthalate

Concen: 0.45 ng

RT: 21.66 min Scan# 1371

Delta R.T. 0.02 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Instrument :

BNA_E





ClientSampleId :

SSTDCCC0.4EC

Tot Ion:	149	Resp:	2938
Ion Ratio	Lower	Upper	
149	100		
167	28.7	22.1	33.1
279	0.0	0.0	0.0

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM



#30

Indeno(1,2,3-cd)pyrene

Concen: 0.37 ng




RT: 26.65 min Scan# 1719


Delta R.T. 0.00 min

Lab File: BE092274.D

Acq: 19 Aug 2016 17:37

Tgt Ion:	276	Resp:	28051
Ion Ratio	Lower	Upper	
276	100		
138	26.5	21.3	31.9
277	24.5	19.7	29.5









#31
Pervlene-d12
Concen: 5.00 ng
RT: 24.13 min Scan# 1548
Delta R.T. -0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.5	29.3
265	22.3	18.2	27.4




Manual Integrations
APPROVED


sohil
8/22/2016 3:50:09 PM



#32
Benzo(b)fluoranthene
Concen: 0.38 ng
RT: 23.39 min Scan# 1484
Delta R.T. -0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.0	19.0	28.4
125	12.5	10.3	15.5









#33
Benzo(k)fluoranthene
Concen: 0.41 ng
RT: 23.44 min Scan# 1488
Delta R.T. -0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.9	19.0	28.6
125	12.6	10.6	15.8




Manual Integrations
APPROVED


sohil
8/22/2016 3:50:09 PM



#34
Benzo(a)pyrene
Concen: 0.38 ng
RT: 24.02 min Scan# 1538
Delta R.T. -0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.4	20.3	30.5
125	14.4	11.8	17.6









#35
Dibenzo(a,h)anthracene
Concen: 0.37 ng
RT: 26.67 min Scan# 1720
Delta R.T. 0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion	Ratio	Lower	Upper
278	100		
139	20.0	16.2	24.2
279	27.0	21.4	32.2

Manual Integrations
APPROVED

sohil
8/22/2016 3:50:09 PM



#36
Benzo(g,h,i)perylene
Concen: 0.39 ng
RT: 27.42 min Scan# 1764
Delta R.T. 0.00 min
Lab File: BE092274.D
Acq: 19 Aug 2016 17:37

Tgt Ion	Ratio	Lower	Upper
276	100		
277	26.0	21.4	32.2
138	24.3	19.4	29.2

