

Data Path : U:\HPCHEM1\BNA E\Data\BE082217\
 Data File : BE093879.D
 Acq On : 22 Aug 2017 17:16
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Aug 23 01:59:04 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 07 14:30:21 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.603	0.480	20.4	65	-0.01
3	n-Nitrosodimethylamine	0.625	0.557	10.9	71	0.01
4 S	2-Fluorophenol	1.294	1.116	13.8	72	0.00
5 S	Phenol-d6	1.976	1.711	13.4	74	0.01
6	bis(2-Chloroethyl)ether	1.530	1.318	13.9	73	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
8 S	Nitrobenzene-d5	0.342	0.294	14.0	79	0.00
9	Naphthalene	4.587	4.027	12.2	77	0.00
10	Hexachlorobutadiene	0.170	0.143	15.9	74	0.00
11 SURR	2-Methylnaphthalene-d10	0.643	0.564	12.3	77	0.00
12	2-Methylnaphthalene	0.774	0.668	13.7	76	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
14 S	2,4,6-Tribromophenol	0.110	0.082	25.5#	71	0.03
15 S	2-Fluorobiphenyl	1.577	1.379	12.6	75	0.00
16	Acenaphthylene	2.011	1.740	13.5	77	0.00
17	Acenaphthene	1.355	1.160	14.4	73	0.00
18	Fluorene	1.752	1.484	15.3	73	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	68	0.00
20	4-Bromophenyl-phenylether	0.113	0.151	-33.6#	92	0.00
21	Hexachlorobenzene	0.121	0.140	-15.7	76	0.00
22	Pentachlorophenol	0.084	0.051	39.3#	44#	0.03
23	Phenanthrene	0.852	0.887	-4.1	72	0.03
24	Anthracene	1.294	1.006	22.3	54	0.00
25 SURR	Fluoranthene-d10	4.564	4.528	0.8	64	0.00
26	Fluoranthene	1.397	1.353	3.1	63	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	68	0.01
28	Pvrene	1.310	1.158	11.6	61	0.00
29 S	Terphenyl-d14	0.713	0.637	10.7	62	0.00
30	Benzo(a)anthracene	1.269	1.047	17.5	58	0.00
31	Chrysene	1.295	1.149	11.3	61	0.00
32	Bis(2-ethylhexyl)phthalate	2.721	2.238	17.8	55	0.00
33	Indeno(1,2,3-cd)pyrene	1.248	0.822	34.1#	45#	0.05
34 I	Perylene-d12	1.000	1.000	0.0	66	0.02
35	Benzo(b)fluoranthene	1.387	1.065	23.2	50	0.01
36	Benzo(k)fluoranthene	1.386	1.279	7.7	62	0.01
37 C	Benzo(a)pyrene	1.309	1.095	16.3	55	0.02
38	Dibenzo(a,h)anthracene	1.225	0.841	31.3#	46#	0.04
39	Benzo(a,h,i)perylene	1.229	0.871	29.1#	47#	0.03

Data Path : U:\HPCHEM1\BNA E\Data\BE082217\
Data File : BE093879.D
Acq On : 22 Aug 2017 17:16
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Aug 23 01:59:04 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE080717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 07 14:30:21 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			