

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE082218\
 Data File : BE097352.D
 Acq On : 22 Aug 2018 18:01
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE082218

Quant Time: Aug 22 18:41:08 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE082218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 22 18:07:26 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	109	0.00
2	1,4-Dioxane	0.400	0.431	-7.7	121	0.00
3	n-Nitrosodimethylamine	0.400	0.419	-4.7	121	0.00
4 S	2-Fluorophenol	0.400	0.381	4.8	116	-0.01
5 S	Phenol-d6	0.400	0.368	8.0	112	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.396	1.0	116	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	0.00
8 S	Nitrobenzene-d5	0.400	0.385	3.8	117	0.00
9	Naphthalene	0.400	0.405	-1.3	116	0.00
10	Hexachlorobutadiene	0.400	0.412	-3.0	118	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.394	1.5	115	0.00
12	2-Methylnaphthalene	0.400	0.401	-0.3	117	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.400	0.382	4.5	122	0.00
15 S	2-Fluorobiphenyl	0.400	0.423	-5.7	118	0.00
16	Acenaphthylene	0.400	0.405	-1.3	120	0.00
17	Acenaphthene	0.400	0.411	-2.7	118	0.00
18	Fluorene	0.400	0.417	-4.2	120	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	107	0.00
20	4-Bromophenyl-phenylether	0.400	0.412	-3.0	123	0.00
21	Hexachlorobenzene	0.400	0.412	-3.0	119	0.00
22	Pentachlorophenol	0.400	0.445	-11.2	122	-0.01
23	Phenanthrene	0.400	0.412	-3.0	120	0.00
24	Anthracene	0.400	0.405	-1.3	121	0.00
25 SURR	Fluoranthene-d10	0.400	0.393	1.8	116	0.00
26	Fluoranthene	0.400	0.404	-1.0	119	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	106	0.00
28	Pvrene	0.400	0.441	-10.2	126	0.00
29 S	Terphenyl-d14	0.400	0.425	-6.2	121	0.00
30	Benzo(a)anthracene	0.400	0.400	0.0	116	0.00
31	Chrysene	0.400	0.407	-1.7	115	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.389	2.8	111	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.352	12.0	106	0.00
34 I	Perylene-d12	0.400	0.400	0.0	104	0.00
35	Benzo(b)fluoranthene	0.400	0.392	2.0	113	0.00
36	Benzo(k)fluoranthene	0.400	0.402	-0.5	115	0.00
37 C	Benzo(a)pyrene	0.400	0.373	6.8	112	0.00
38	Dibenzo(a,h)anthracene	0.400	0.362	9.5	106	0.00
39	Benzo(a,h,i)perylene	0.400	0.356	11.0	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE082218\
Data File : BE097352.D
Acq On : 22 Aug 2018 18:01
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE082218

Quant Time: Aug 22 18:41:08 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE082218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 22 18:07:26 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				